

Washington University School of Medicine

Digital Commons@Becker

Open Access Publications

2010

Image-based calibration of a deformable mirror in wide-field microscopy

Diwakar Turaga

Washington University School of Medicine in St. Louis

Timothy E. Holy

Washington University School of Medicine in St. Louis

Follow this and additional works at: https://digitalcommons.wustl.edu/open_access_pubs

Please let us know how this document benefits you.

Recommended Citation

Turaga, Diwakar and Holy, Timothy E., "Image-based calibration of a deformable mirror in wide-field microscopy." *Applied Optics*. 49, 11. 2030-2040. (2010).

https://digitalcommons.wustl.edu/open_access_pubs/3563

This Open Access Publication is brought to you for free and open access by Digital Commons@Becker. It has been accepted for inclusion in Open Access Publications by an authorized administrator of Digital Commons@Becker. For more information, please contact vanam@wustl.edu.

Image-based calibration of a deformable mirror in wide-field microscopy

Diwakar Turaga and Timothy E. Holy*

Department of Anatomy and Neurobiology, Washington University School of Medicine, St. Louis Missouri 63110, USA

*Corresponding author: holy@wustl.edu

Received 22 December 2009; revised 1 March 2010; accepted 5 March 2010;
posted 9 March 2010 (Doc. ID 121826); published 6 April 2010

Optical aberrations limit resolution in biological tissues, and their influence is particularly large for promising techniques such as light-sheet microscopy. In principle, image quality might be improved by adaptive optics (AO), in which aberrations are corrected by using a deformable mirror (DM). To implement AO in microscopy, one requires a method to measure wavefront aberrations, but the most commonly used methods have limitations for samples lacking point-source emitters. Here we implement an image-based wavefront-sensing technique, a variant of generalized phase-diverse imaging called multiframe blind deconvolution, and exploit it to calibrate a DM in a light-sheet microscope. We describe two methods of parameterizing the influence of the DM on aberrations: a traditional Zernike expansion requiring 1040 parameters, and a direct physical model of the DM requiring just 8 or 110 parameters. By randomizing voltages on all actuators, we show that the Zernike expansion successfully predicts wavefronts to an accuracy of approximately 30 nm (rms) even for large aberrations. We thus show that image-based wavefront sensing, which requires no additional optical equipment, allows a simple but powerful method to calibrate a deformable optical element in a microscope setting. © 2010 Optical Society of America

OCIS codes: 110.1080, 180.2520.

1. Introduction

Light is refracted by biological tissues. This interaction can be exploited to generate image contrast; however, refractions also present a significant hindrance to image resolution deeper into tissue. The interactions between light and tissue are conventionally discussed in terms of two extremes: “scattering” typically describes the effects of small inhomogeneities in tissue, whereas the term “aberrations” most commonly refers to refractions induced by bulk (average) properties of tissue [1]. One common source for aberrations is the mismatch in index of refraction between the immersion fluid and the sample; for example, water or saline has a refractive index near 1.33, but tissue typically has a variable refractive index ranging from 1.36 to 1.40 [2].

In a situation in which the tissue is face-on with the objective (i.e., the tissue surface is orthogonal to objective’s optical axis), this refractive index mismatch leads primarily to two types of aberration, defocus and spherical aberration [3]. The defocus aberration is often not even noticed (it is corrected by changing the focus of the objective), but the spherical aberration typically remains uncorrected and serves as an impediment to imaging deeper into the tissue. More problematic is the case where the tissue is not perfectly flat and/or the tissue surface is not orthogonal to the objective axis, because additional aberrations, which can be substantially larger, are introduced into the images [4]. An extreme case of such tilted imaging is found in a light-sheet-based microscopic technique called objective coupled planar illumination microscopy (OCPI) [5].

In typical light-sheet microscopy (also sometimes called planar illumination microscopy) a cylindrical lens is used to create a sheet of light [5–8]. This sheet

of light is placed at the focal plane of the objective, and the tissue is placed in this overlap region. This arrangement allows only the in-focus region of the tissue to be illuminated and allows the entire illuminated plane to be imaged simultaneously. Thus light-sheet microscopy allows high-speed and low-phototoxic imaging. To use such light-sheet microscopy to image large samples (i.e., the surface of the mouse brain *in vivo*), one should minimize the distance in tissue traversed by the excitation light and emitted light; consequently, the light sheet is tilted with respect to the tissue surface, and the objective is tilted correspondingly [Fig. 1(a)]. This tilted imaging introduces sizable new optical aberrations, including defocus, coma, and astigmatism. Previously we showed that the defocus aberration can be corrected by tilting the angle of the light sheet by a few degrees [4]. We hypothesize that wavefront-correcting techniques such as adaptive optics (AO) can be used to correct the remaining aberrations.

AO has been used in a variety of settings, initially by astronomers to correct the loss of resolution in images taken from Earth-bound telescopes due to atmospheric turbulence [9]. In AO, a wavefront sensor (typically a Shack–Hartmann wavefront sensor, SHWFS) is used to measure the aberrations, and a deformable mirror (DM) is used to correct the wavefront to achieve diffraction-limited imaging. The resulting improvement in image quality has made AO an integral part of major new ground-based telescope designs. Similar AO systems have also been used in vision science, where the aberrations of the eye can be measured and corrected [10].

The most crucial element of any AO system is the wavefront sensor; here we focus on the associated problem as it applies to wide-field microscopy. A traditional SHWFS is difficult to apply directly to most applications in microscopy because tissue samples typically lack point-source emitters [3]; generalizations to extended samples have been described [11], and this might be a useful approach for microscopy,

but the need to sample the wavefront at high resolution means that aberrations can be measured only over a region corresponding to a few tens of pixels. Several alternative methods for wavefront sensing have been developed. Coherence-gated techniques [12,13] are applicable only for scattered light, and these methods introduce some degree of complexity into the apparatus. A more general and (in terms of instrumentation) simpler approach is to use the images themselves to estimate the wavefront aberration. Approaches that iteratively improve the sharpness of an image have been found to be applicable in microscopy [14,15]. But such methods require large number of iterations and/or a large number of images and thus place constraints on the speed and fluorescence levels of the biological preparations. Phase retrieval is a method used to measure the point-spread-function (PSF) of the microscope [16], but such a method is not capable of measuring wavefront aberrations of extended objects. However, a closely related method—phase-diverse imaging (PDI)—is capable of measuring wavefront aberrations of extended objects [17,18]. PDI can be described as follows: in a typical imaging experiment there are two unknowns, (i) the object and (ii) the wavefront aberration; a single image is insufficient to accurately measure the two unknowns. In typical PDI, one image is acquired with the camera in focus, and a second image is acquired with the camera slightly out of focus. Thus with the two known images the two unknowns (object and wavefront aberration) can be extracted computationally [17,18]. In principle, such an image-based wavefront sensor is readily applicable in the high-signal-to-noise ratio imaging performed through OCPI microscopy.

In order to apply AO to OCPI microscopy, a DM needs to be placed in the light path of the microscope; the DM has a number of control signals, and the effect of these control signals on the aberration structure must be calibrated. SHWFS or interferometry-based systems can be used to calibrate a DM, but




Fig. 1. (Color online) AO-OCPI schematic: (a) Experimental setup for AO-OCPI microscope. A DM is placed behind the back aperture of the objective. The light reflected off the DM is imaged onto a camera. (b), (c) Schematic of wavefront aberration when the DM is flat and when one actuator on the DM is moved.

that would entail adding a new optical system to the microscope. Since this problem is conceptually equivalent to the wavefront sensing needed for AO, PDI algorithms should also be useful for calibration [19] [Fig. 1(b)]. In practice we apply a technique known as multiframe blind deconvolution (MFBD) [20]. In MFBD one usually acquires multiple images of the same object with multiple unknown aberrations (here caused by DM actuator movement) and exploits the fact that the object is constant to infer the structure of the aberrations. MFBD differs from PDI only by having fewer constraints on the aberrations. Thus, virtually all of the mathematical and algorithmic apparatus can be shared both for aberration correction and DM calibration. In this paper we describe the calibration of a DM in an OCPI microscopy setting. It should be noted that such image-based wavefront sensing is not limited to light-sheet microscopy, but can be implemented in other imaging applications.

2. AO-OCPI Optical Layout

In OCPI microscopy the optics needed to form a light sheet are rigidly coupled to the objective, illuminating just the focal plane of the objective [5] [Fig. 1(a)]. To permit imaging of extended neural tissues, we tilt the objective (with the coupled laser sheet) from the traditional face-on imaging to an angle of 30°. This minimizes the distance traveled through tissue by both the excitation light and the emitted light.

To correct the resulting aberrations, a DM (Mirao 52-d, Imagine Optics) is placed behind the back aperture of the microscope objective (20× infinity-corrected, 0.5 NA, water immersion, Olympus). The light collected by the objective is reflected off the DM before it is focused onto a camera (GRAS-14S5M, Point Grey) by a 200 mm tube lens (Edmund Optics) [Fig. 1(a)].

The Mirao 52-d is a 52-actuator DM [21], with the actuators encompassing a circular pupil of 15 mm diameter [Fig. 2(a)]. The reflective surface is made of a silver-coated sheet with an array of permanent magnets on the back side; to change the shape of the mirror, one applies force by controlling the current in coils placed opposite each magnet [Fig. 2(b)].




Fig. 2. (Color online) Mirao 52-d. (a) Schematic image of the Mirao 52-d DM. The 52 actuators encompass a pupil of diameter 15 mm. The numbering of the actuators presented here is used in rest of the paper. (b) The mirror is made of a sheet of mirror with voltage-controlled magnetic actuators on the back surface of the mirror.

3. Phase-Diverse Imaging: Theory

Consider a base object emitting light with (scalar) intensity $f(\mathbf{x})$ at position \mathbf{x} in a two-dimensional plane. (One challenge in applying traditional PDI to microscopy is the extended, three-dimensional nature of typical objects. The localization of excitation in light-sheet microscopy makes its application more straightforward.) A total of K different images are collected of this fixed object; these images have different aberrations, which here are generated by different settings of the voltages for the DM. These K diversity images are denoted $d_k(\mathbf{x})$, $k = 1, \dots, K$. The PSF for the k th diversity image is denoted $s_k(\mathbf{x})$, and because the imaging path is incoherent we write

$$s_k(\mathbf{x}) = |h_k(\mathbf{x})|^2, \quad (1)$$

where h_k is the inverse Fourier transform of the optical transfer function for coherent illumination, H_k . Explicitly,

$$h_k(\mathbf{x}) = \text{FT}^{-1}[H_k(\mathbf{u})] = \int d\mathbf{u} e^{-2\pi i \mathbf{u} \cdot \mathbf{x}} H_0(\mathbf{u}) e^{i\phi_k(\mathbf{u})}, \quad (2)$$

where FT^{-1} is the inverse Fourier transform, \mathbf{u} is a point in the pupil, H_0 is the aperture mask (usually zero outside the pupil and one inside) and ϕ_k is the aberration phase associated with the k th diversity image.

The unknowns, f and the set $\{\phi_k\}$ of all aberration phases, will be determined by nonlinear optimization, minimizing the square difference between observed (d_k) and predicted ($s_k * f$, where $*$ is convolution) images:

$$E[f, \{\phi_k\}] = \sum_{k=1}^K \int d\mathbf{u} |D_k(\mathbf{u}) - F(\mathbf{u})S_k(\mathbf{u})|^2, \quad (3)$$

where D_k , F , S_k are the Fourier transforms of d_k , f , s_k , respectively.

As shown previously [17,18], the penalty, Eq. (3), can be converted to a pure function of ϕ_k by substituting the analytic solution for the optimum F ,

$$F = \sum_{k=1}^K D_k S_k^\dagger / \sum_{l=1}^K |S_l|^2, \quad (4)$$

where \dagger is the complex conjugate, leading to a phase-only penalty function

$$E'[\{\phi_k\}] = \int d\mathbf{u} \frac{\left| \sum_k D_k(\mathbf{u}) S_k^\dagger(\mathbf{u}) \right|^2}{\sum_l |S_l(\mathbf{u})|^2} - \int d\mathbf{u} \sum_k |D_k(\mathbf{u})|^2. \quad (5)$$

We note that the second term does not depend upon the aberration. In minimizing the resulting penalty function, the speed of convergence is substantially enhanced by using the penalty gradient [18], which for our parameterization is

$$\frac{\delta E'}{\delta \phi_k(\mathbf{u})} = 4 \operatorname{Im} \left[\sum_{k=1}^K H_k(\mathbf{u}) (Z_k * H_k^\dagger)(\mathbf{u}) \right], \quad (6)$$

where

$$Z_k = \left[\sum_l |S_l|^2 \left(\sum_j D_j S_j^\dagger \right) D_k^\dagger - \left| \sum_j D_j S_j^\dagger \right|^2 S_k^\dagger \right] / \left(\sum_l |S_l|^2 \right)^2. \quad (7)$$

From Eqs. (1) and (2), note that the images are unchanged by adding a constant to ϕ_k (a “piston shift”). Thus, the mean value of ϕ_k is not meaningful. Moreover, the images are also unaltered by the replacement $\phi(\mathbf{u}) \rightarrow -\phi(-\mathbf{u})$, because this merely results in $h(\mathbf{x})$'s being replaced with its complex conjugate. In our calibration procedure, we used the knowledge that the actuators are arranged in a grid to ensure that a consistent sign convention was adopted for all actuators.

4. Parameterizations of the Aberration Function ϕ

For a parametric representation of $\{\phi_k\}$, Eq. (6) allows the derivative with respect to any parameter p to be calculated via the chain rule,

$$\frac{\partial}{\partial p} = \sum_k \int d\mathbf{u} \frac{\partial \phi_k(\mathbf{u})}{\partial p} \frac{\delta}{\delta \phi_k(\mathbf{u})}. \quad (8)$$

A. Gaussian Parameterization

Since individual actuators are expected to have a localized influence in the pupil plane, one simple (few-parameter) approximation of the aberration phase would be a Gaussian. (Here, the primary value of a Gaussian parameterization is to provide an initial guess for a more accurate method requiring more parameters.) We postulate that the phase aberration can be approximated as

$$\phi(\mathbf{u}) = A e^{-(\mathbf{u}-\mathbf{u}_0)^2/2\sigma^2}, \quad (9)$$

where A is the amplitude of deformation, \mathbf{u}_0 is the center of deformation, and σ is the estimate of the width of the deformation. The partial derivative of ϕ_k with respect to the parameters is

$$\frac{\partial \phi_k}{\partial A} = e^{-(\mathbf{u}-\mathbf{u}_0)^2/2\sigma^2}, \quad (10)$$

$$\nabla_{\mathbf{u}_0} \phi_k = \frac{A(\mathbf{u}-\mathbf{u}_0)}{\sigma^2} e^{-(\mathbf{u}-\mathbf{u}_0)^2/2\sigma^2}, \quad (11)$$

$$\frac{\partial \phi_k}{\partial \sigma} = \frac{A(\mathbf{u}-\mathbf{u}_0)^2}{\sigma^3} e^{-(\mathbf{u}-\mathbf{u}_0)^2/2\sigma^2}. \quad (12)$$

Together with Eqs. (6) and (8), these equations allow one to calculate the gradient of the penalty with respect to the parameters of the Gaussian.

B. Zernike Parameterization

The most common way to parameterize aberrations is with a Zernike expansion [22],

$$\phi_k = \sum_{j=1}^J \alpha_j Z_j, \quad (13)$$

where Z_j is the j th Zernike function (using a single-indexing scheme [23]) and α_j is the coefficient for the j th Zernike function.

The partial derivative of ϕ_k with respect to α_j is simply

$$\frac{\partial \phi_k}{\partial \alpha_j} = Z_j. \quad (14)$$

C. Biharmonic Parameterization

An attractive alternative to these phenomenological parameterizations is to consider a direct physical model of the DM, i.e., parameterizing ϕ in terms of the shape of the surface of the mirror. In the general case, the alignment of the pupil with respect to the DM is unknown, and in any event the projection of the pupil onto the DM surface is compressed along one axis owing to the angle of reflection off the surface (Fig. 1). These statements may be encapsulated as

$$\phi_k(\mathbf{u}) = \psi_k(\mathbf{A}\mathbf{u} + \boldsymbol{\xi}_0), \quad (15)$$

where the matrix \mathbf{A} and offset $\boldsymbol{\xi}_0$ correspond to a rigid body deformation of the pupil, and $\psi_k(\boldsymbol{\xi})$ is effectively the mirror shape represented in two-dimensional coordinates $\boldsymbol{\xi}$ in the plane of the mirror. This representation thus casts the problem as fundamentally a rigid registration problem, but with respect to pupil coordinates (phase) rather than a registration of two

observed images. \mathbf{A} needs to allow rotation and scaling but not shearing ($\mathbf{A} = \mathbf{R}\mathbf{S}\mathbf{R}^T$, where \mathbf{R} is a rotation matrix and \mathbf{S} is diagonal), and thus \mathbf{A} is a generic 2×2 symmetric matrix

$$\mathbf{A} = \begin{pmatrix} a_1 & a_2 \\ a_2 & a_3 \end{pmatrix}. \quad (16)$$

After fitting, the rotation angle and scaling diagonals may be extracted from a singular value decomposition of \mathbf{A} ; in particular, the absolute value of the ratio of the diagonals of \mathbf{S} should be equal to the cosine of the angle of reflection off the mirror surface.

The form of ψ depends on the DM. For the Mirao 52-d, the linearity with respect to perturbation of individual actuators (see Fig. 5 below) suggests a model

$$\psi_k(\xi) = \sum_i (m_i v_{ki} + \zeta_i) b_i(\xi), \quad (17)$$

where v_{ki} is the control voltage applied to the i th actuator in the k th diversity image, m_i and ζ_i are the slope and offset for this actuator, respectively, and $b_i(\xi)$ describes the shape of the surface induced by applying unit voltage to the i th actuator; b_i encapsulates the physics of the device, and in this case the necessary details (equation of motion and boundary conditions) are not publicly disclosed by the manufacturer. However, because this mirror is constructed from an elastic membrane, we postulate that the membrane energy is a function of the curvature, i.e.,

$$E_{\text{membrane}}[b] \propto \int d\xi (\nabla^2 b)^2, \quad (18)$$

where ∇^2 is the Laplacian with respect to coordinates ξ . Hence b satisfies the fourth-order *biharmonic equation*,

$$(\nabla^2)^2 b = 0. \quad (19)$$

This equation of motion needs to be supplemented by the boundary conditions, for which we will assume that the membrane is clamped at some radius R and that this clamp sets both the membrane height and slope to zero on the boundary ($b = 0$ and $\hat{\mathbf{r}} \cdot \nabla b = 0$ on the boundary, where $\hat{\mathbf{r}}$ is the unit radial vector). Consequently, application of a unit force at point \mathbf{c} will induce a (normalized) displacement given by the Green's function [24],

$$b(\tilde{\xi}|\tilde{\mathbf{c}}) = \frac{1}{R^2} |\tilde{\xi} - \tilde{\mathbf{c}}|^2 \log \left(\frac{R^2 |\tilde{\xi} - \tilde{\mathbf{c}}|^2}{|R^2 - \tilde{\mathbf{c}}^\dagger \tilde{\xi}|^2} \right) + \frac{1}{R^4} (R^2 - |\tilde{\xi}|^2)(R^2 - |\tilde{\mathbf{c}}|^2), \quad (20)$$

where $\tilde{\xi} = \xi_x + i\xi_y$ is the complex number formed from the x and y coordinates of ξ and similarly for $\tilde{\mathbf{c}}$. For each actuator, the corresponding \mathbf{c}_i is specified

from the known grid arrangement of the actuators [Fig. 2(a)].

The advantage of this approach is in the comparatively small number of parameters required: rather than needing the first five orders of Zernike functions (20 Zernike coefficients for each actuator, a total of 1040 parameters), here each actuator is represented by only two parameters, m_i and ζ_i . Indeed, in a first-pass optimization one may take $\zeta_i = 0$ and use a common value $m_i = m$ for all actuators (assuming that each electromagnet produces similar force), and thus all 52 actuators contribute just a single parameter, m . One also must fit R and the rigid-deformation parameters \mathbf{A} and ξ_0 , and for calibration with a single bead we also modify Eq. (15) to allow a defocus,

$$\phi_k(\mathbf{u}) = \psi_k(\mathbf{A}\mathbf{u} + \xi_0) + \alpha(2\mathbf{u}^2 - 1), \quad (21)$$

to include the possibility that the bead is slightly above or below the focal plane (which may not be readily detectable in the mirror flat condition but can nevertheless have a substantial effect on the aberrated images). Consequently, this approach requires a total of either 8 parameters (\mathbf{A} , ξ_0 , α , and m) or 110 parameters (\mathbf{A} , ξ_0 , α , and susceptibility and offset, m_i and ζ_i , respectively, for each actuator).

Optimization of these parameters greatly benefits from an analytic calculation of the gradients; for reasons of space we do not present explicit formulas, but their derivation from Eqs. (21) and (20) is entirely straightforward (if slightly tedious). The initial guess for the eight-parameter model is supplied by the user with the help of a custom GUI program, to ensure that the pupil registration parameters do not become trapped in a local minimum far from the optimum solution.

5. Experiments

We imaged a $0.2 \mu\text{m}$ (diffraction-limited) fluorescent bead embedded at the surface of a flat slab of polydimethylsiloxane (PDMS, Dow Corning, DC 184-A and DC 184-B with a weight ratio of 10:1). Each actuator was manipulated by applying voltages ranging from -0.1 to 0.1 V with 0.01 V steps (all the other actuators were maintained at 0 V), and an image was collected at each step. Thus, we obtained 21 images for each of the 52 actuators, for a total of 1092 images. In Fig. 3 we show the images obtained by moving one of the actuators [actuator 22, see Fig. 2(a)] from -0.09 to 0.1 V (data from all actuators is supplied as Media 1).

6. Results

A. Calibration of DM Using Zernike Parameterization

Each actuator was calibrated separately. The image-based wavefront reconstruction described in Section 3 was applied on the 21 images—obtained from applying different voltages to a single actuator—to calculate the unknown underlying aberration function ϕ .




Fig. 3. Images of a $0.2\mu\text{m}$ bead obtained after applying from -0.09 to 0.1 V to actuator 22 [see Fig. 2(a) for the location of the actuator]. Each image is 64×64 pixels, with $1\text{ pixel} = 0.29\mu\text{m} \times 0.29\mu\text{m}$. (All consequent bead images have the same dimensions). Media 1 shows images obtained from moving each of the 52 actuators.

In the optimization, the initial estimate of ϕ was created by fitting a Gaussian model of the aberration, which (having far fewer parameters) allowed rapid and relatively exhaustive search. In Fig. 4 we show that the images obtained by using Gaussian parameterization of ϕ yielded moderate agreement between experimental and calculated images. We expanded the ϕ obtained from the Gaussian parameterization in terms of the first 5 orders of Zernike functions (20 parameters in total per actuator) and used the obtained Zernike coefficients as the initial guess for the Zernike-based optimization. We found that the calculated estimate of ϕ did not improve substantially when a larger number of Zernike functions were used (see Fig. 12 below).

In Fig. 5(a) we show the optimized values for the Zernike coefficients. Each of the Zernike parameters varies approximately linearly with voltage; the slope of the relationship is plotted as a phase plot in Fig. 5(b). The single-hump peak depicts the movement of an actuator at the location of the peak. In Fig. 6 we demonstrate the accuracy of the calibration of one actuator [number 22, Fig. 2(a)] by showing the




Fig. 4. Measured (top) and calculated (bottom) images of a bead when different voltages are applied to actuator 22. The calculated images were obtained from the calculated estimate of ϕ using Gaussian parameterization.

measured and calculated images for images obtained when different voltages are applied to this actuator.

Each actuator was calibrated individually with an identical voltage series comprising 21 images. One observes that the resulting aberrations define a grid structure reminiscent, as one might expect, of the underlying grid of actuators (Fig. 7). To visualize the alignment between Figs. 2(a) and 7 one needs to reflect the DM actuator grid across the vertical. In Media 2 we show the measured and calculated images obtained for all the actuators.

In Fig. 5(a) we can see that the Zernike coefficients have a nonzero component at zero applied voltage. This yields an estimate of the baseline aberrations of the system, which include any imperfections in the system optics (which tend to be small) and deviations in the mirror shape from nominally flat (which can be more substantial). In Fig. 8(a) we show that the zero offsets measured for each of the actuators have similar values even though they were measured independently. By calculating the means of the offsets for each of the 20 Zernike coefficients, we estimated the shape of the mirror when all actuators are set to zero voltage [Fig. 8(b)]. The flatness of the mirror can be improved by supplying an array of offset voltages \mathbf{v}_0 , with

$$\mathbf{v}_0 = -\mathbf{M}^+ \mathbf{Z}_0, \quad (22)$$

where \mathbf{Z}_0 is the vector of Zernike coefficients in the nominally flat condition, \mathbf{M} is the matrix formed from the slope of the Zernike coefficients for each actuator [Fig. 5(a)], and $+$ denotes the Moore–Penrose pseudoinverse.

B. Calibration of DM Using a Physical Membrane-Based Model of DM

We performed the 8- and 110-parameter biharmonic parameterizations of the aberrations caused by individual actuators. The eight-parameter model was obtained by using all 21×52 images to fit the parameters \mathbf{A} , ξ_0 , α , and a single value of m for all actuators, while setting $\zeta = 0$. The resulting phases for each of the actuators are shown in Fig. 9.

To obtain the actuator phases for the 110-parameter model, \mathbf{A} , ξ_0 and α were held constant while m and ζ were fitted separately for each actuator. In Fig. 10 we show the value of m_i obtained for each of the actuators. The m_i values appear to be dependent on the location of the corresponding actuator—suggesting an underlying asymmetry in the DM.

In Fig. 11 we show the experimental and calculated images (using biharmonic parameterization) obtained when voltage is applied to actuator 22. The quality of the agreement is lower than observed for the Zernike fit (Fig. 6). Therefore, henceforth we adopted the Zernike parameterization.

To quantitatively compare the different parameterizations implemented in this paper, we calculated the fitting error [given by Eq. (3)] for each of the cases (Fig. 12). Gaussian and biharmonic parameteriza-




Fig. 5. (Color online) Actuator 22 calibration using a Zernike parameterization. (a) Each of the 20 Zernike coefficients varies nearly linearly with voltage. The slope of the relationship is plotted as a phase plot in (b). (b) The single-hump peak of the phase plot depicts the movement of the actuator at the location of the peak (colorbar units are radians per volt).

tions lead to the greatest fitting error. The fitting error decreased as greater numbers of Zernike coefficients were used. Most of the decrease in fitting error was achieved by including Zernike polynomials up through the fifth order, and so we truncated the expansion at this point for the tests described below.

C. Testing the Calibration

The calibration was performed individually for each actuator, yet for AO one must manipulate all of the actuators simultaneously to compensate for arbitrary aberrations. To test whether the calibration procedure suffices in this circumstance, we specified random voltages for all the actuators in the DM and measured the resulting images of beads. Using the Zernike calibration data, we estimated the total wavefront aberration as the linear sum of aberrations caused by individual actuators. This predicted aberration was then used to calculate the theoretical PSF to obtain a calculated estimate of the image. A schematic of the procedure is shown in Fig. 13 (see Media 3). In the figure, the unaberrated image is the image acquired when all the actuators are set to zero voltage. The predicted aberrated image is produced by the convolution of the estimated wavefront aberration with a point (since the underlying object is unknown).




Fig. 6. Measured (top) and calculated (bottom) images of a bead when different voltages are applied to actuator 22. The calculated images were obtained from the Zernike parameterization. Media 2 shows corresponding images obtained for all actuators.

To explore whether the deviation from nominal flatness requires compensation, we considered the random voltage test both before and after correcting for the true mirror flatness (Fig. 14). The calculated images are modestly improved by including the flatness correction. We performed such a random-voltage test, using 50 different sets of random voltages, with all of them leading to good agreement between the measured and the calculated images (Fig. 14, Media 3).

The accuracy of the calibration can also be tested by comparing the calibration-derived random aberration to the true random aberration. We calculated the true random aberration by using PDI, as described in Section 3. We used the unaberrated image and the randomly aberrated image as the two




Fig. 7. (Color online) Phase plots obtained for all 52 actuators, using Zernike parameterization. Each of the phase plots is scaled independently to demonstrate the underlying differences.




Fig. 8. (Color online) DM flat. (a) Magnitude of Zernike coefficients for each of the actuators at zero applied voltage obtained from optimization of each actuator independently; note the consistency of the fitting result. The first two Zernike coefficients contribute to overall tip and tilt of the PSF and are not shown in this figure. (b) Phase present at zero applied voltage (colorbar units are radians).

input images and optimized the phase. As an initial estimate we used the calibration-derived phase, which consisted of up to fifth-order Zernike coefficients. To explore a possible role for the higher-order coefficients in representing the error, we extended the phase to consist of up to sixth-order Zernike coefficient, setting the sixth-order Zernike coefficients initially to zero. The Zernike coefficients were optimized to produce the optimum aberrated phase. Figure 15 shows the calibration-derived phase (predicted phase), the PDI-derived phase (optimized phase) and the difference in the two phases for the images shown in Fig. 13. Figure 15(d) shows that the calibration-derived estimates for the Zernike coefficients are very close to the optimized values. We performed the same calculation on each of the




Fig. 9. (Color online) Phase plot obtained for all 52 actuators, using the 8-parameter biharmonic parameterization. Each of the phase plots is scaled independently to demonstrate the underlying differences.




Fig. 10. (Color online) Values of m_i obtained for each of the actuators after a 110 parameter biharmonic parameterization of ϕ (colorbar units are radians per volt).

50 sets of the random voltages tested (Media 3) and plotted the rms error in Fig. 15(e). The rms error averaged across realizations was ~ 30 nm, which is below the threshold for making a significant impact on image quality. Thus we conclude that the calibrated DM data yields an accurate model of the mirror.

7. Discussion

By using only an image-based analysis we have shown that MFBD can be used to accurately calibrate a DM. The decision whether to use MFBD or other methods depends on several factors. (i) Equipment: MFBD-based calibration does not require specialized optical equipment such as SHWFS and interferometers. The MFBD method even avoids the need for a beam splitter [19]. (ii) Speed: the speed of wavefront sensing in SHWFS and interferometry—by virtue of being specialized devices to measure wavefronts—is nearly instantaneous. In




Fig. 11. Measured (top) and calculated (middle, bottom) images of a bead when different voltages are applied to actuator 22. The calculated images were obtained by using a biharmonic parameterization with 8 (middle) and 110 (bottom) parameters. Media 4 shows corresponding images obtained for all actuators.




Fig. 12. (Color online) Comparison of fitting errors: (a) fitting error between experimental and calculated images using different ϕ parameterizations, for actuator 22. The different parameterizations are Gaussian (G), from second- to eighth-order Zernike parameters, bi-harmonic parameterization using 8 parameters (B8), and biharmonic parameterization using 110 parameters (B110). (b) Total fitting error between the experimental and calculated images for all actuators.

MFBD-based calibration, the wavefront calculation is a computationally intensive process. The calibration of a single actuator, using 21 images (see Section 6) took ~ 3 min (this includes a preliminary GUI-based Gaussian estimate, followed by a Zernike-based optimization of the aberration function ϕ) on a 32 Gbyte RAM, single-core processor. Thus the calibration of the 52 actuator DM took ~ 2.5 h. Fortunately, such a CPU-intensive calculation

needs to be performed only once, during the construction of the microscope. (iii) Accuracy: the MFBD-based calibration is ultimately limited by the noise in the images. We have shown that the method produces an error of ~ 30 nm, which is well below the threshold where the error has a significant impact on the image quality. We note that this accuracy compares well with previous studies [19], in that this represents the *open-loop* calibration error, and that it is




Fig. 13. (Color online) A random set of voltages are applied to the DM to produce the acquired image. The Zernike-parameterization-based calibration of the DM is used to calculate the predicted phase produced by the set of random voltages (colorbar units are radians). The predicted phase is then used to create the predicted image. The agreement between the acquired and predicted images demonstrates the accuracy of the calibration of the DM. Media 3 shows images obtained from using the 50 sets of random voltages.




Fig. 14. Sets of random voltages were applied to the DM to obtain the experimental (Exp.) images. Zernike-based calibration of the DM was used without (Calc.-) and with (Calc.+) offset correction to obtain the calculated images for the given set of random voltages. Each column represents images obtained from a different set of random voltages.

a small fraction of a much larger underlying aberration (Figs. 12 and 15) of a type—generated by applying random voltages to all actuators—not included in the data used to perform the calibration.

We explored two different parameterizations for the phase aberration. The Zernike parameterization of the DM (using 1040 parameters) led to accurate calibration of the DM. The parameterization based

on a physical model was attractive because of the smaller number of parameters (8 or 110). This reduction in number of parameters required to describe a 52-actuator DM can be attributed to the gridlike placement of the actuators on the DM and the physical characteristics of a membrane mirror. The physical model led to a useful but less-accurate calibration of the DM. We suspect this reflects the fact that the bi-harmonic model is an imprecise description of the underlying physics of the Mirao 52-d, for which important details are not publicly available. For example, if the membrane is under some tension, the equation of motion might require the addition to Eq. (18) of a term penalizing the gradient of b . Nevertheless, our results demonstrate that physical models have considerable potential for parameterizing aberrations with small numbers of parameters. Given that PDI MFBD is fundamentally an optimization technique, reducing the number of parameters may improve the speed and/or robustness of convergence, potentially extending the range of applications of PDI.

MFBD was shown to be an accurate and efficient method for calibrating a DM in a light-sheet microscope. A major advantage of this approach is that the hardware demands are minimal; most of the complexity is contained in the mathematical analysis




Fig. 15. (Color online) (a) Predicted phase from calibration data (colorbar units are radians). (b) Optimized phase calculated by using PDI on the unaberrated and aberrated images shown in Fig. 13. (c) Difference in the two phases. Note that the colormap has been rescaled to show fine detail. (d) Zernike coefficients for predicted (a) and optimized (b) phases. (e) The rms error (in nanometers) between the two phases for all 50 sets of random voltages tested. The red \times represents the rms error obtained for the example shown in (d).

and therefore encapsulated in software. We have made our software freely available at <http://holylab.wustl.edu>. These developments should contribute to more widespread application of AO in microscopy.

The authors thank Zhongsheng Guo for computational help. The research was supported by the McKnight Technological Innovations in Neuroscience Award and the National Institutes of Health (NINDS/NIAAA R01NS068409, T. E. Holy).

References

1. J. B. Pawley, *Handbook of Biological Confocal Microscopy*, 3rd ed. (Springer, 2006).
2. J. J. J. Dirckx, L. C. Kuypers, and W. F. Decraemer, "Refractive index of tissue measured with confocal microscopy," *J. Biomed. Opt.* **10**, 044014 (2005).
3. M. J. Booth, "Adaptive optics in microscopy," *Philos. Transact. R. Soc. A* **365**, 2829–2843 (2007).
4. D. Turaga and T. E. Holy, "Miniaturization and defocus correction for objective-coupled planar illumination microscopy," *Opt. Lett.* **33**, 2302–2304 (2008).
5. T. F. Holekamp, D. Turaga, and T. E. Holy, "Fast three-dimensional fluorescence imaging of activity in neural populations by objective-coupled planar illumination microscopy," *Neuron* **57**, 661–672 (2008).
6. E. Fuchs, J. S. Jaffe, R. A. Long, and F. Azam, "Thin laser light sheet microscope for microbial oceanography," *Opt. Express* **10**, 145–154 (2002).
7. J. Huisken, J. Swoger, F. Del Bene, J. Wittbrodth, and E. H. K. Stelzer, "Optical sectioning deep inside live embryos by selective plane illumination microscopy," *Science* **305**, 1007–1009 (2004).
8. H.-U. Dodt, U. Leischner, A. Schierloh, N. Jährling, C. P. Mauch, K. Deininger, J. M. Deussing, M. Eder, W. Zieglgänsberger, and K. Becker, "Ultramicroscopy: three-dimensional visualization of neuronal networks in the whole mouse brain," *Nat. Methods* **4**, 331–336 (2007).
9. R. K. Tyson, *Introduction to Adaptive Optics* (SPIE Press, 2000).
10. J. Porter, H. Queener, J. Lin, K. Thorn, and A. Awwal, *Adaptive Optics for Vision Science* (Wiley-Interscience, 2006).
11. L. A. Poyneer, "Scene-based Shack–Hartmann wave-front sensing: analysis and simulation," *Appl. Opt.* **42**, 5807–5815 (2003).
12. M. Feierabend, M. Ruckel, and W. Denk, "Coherence-gated wave-front sensing in strongly scattering samples," *Opt. Lett.* **29**, 2255–2257 (2004).
13. B. Hermann, E. J. Fernandez, A. Unterhuber, H. Sattmann, A. F. Fercher, W. Drexler, P. M. Prieto, and P. Artal, "Adaptive-optics ultrahigh-resolution optical coherence tomography," *Opt. Lett.* **29**, 2142–2144 (2004).
14. M. J. Booth, M. A. A. Neil, R. Juškaitis, and T. Wilson, "Adaptive aberration correction in a confocal microscope," *Proc. Natl. Acad. Sci. USA* **99**, 5788–5792 (2002).
15. D. Debarre, E. J. Botcherby, T. Watanabe, S. Srinivas, M. J. Booth, and T. Wilson, "Image-based adaptive optics for two-photon microscopy," *Opt. Lett.* **34**, 2495–2497 (2009).
16. B. M. Hanser, M. G. L. Gustafsson, D. A. Agard, and J. W. Sedat, "Phase-retrieved pupil functions in wide-field fluorescence microscopy," *J. Microscopy* **216**, 32–48 (2004).
17. R. A. Gonsalves, "Phase retrieval and diversity in adaptive optics," *Opt. Eng.* **21**, 829–832 (1982).
18. R. G. Paxman, T. J. Schulz, and J. R. Fienup, "Joint estimation of object and aberrations by using phase diversity," *J. Opt. Soc. Am. A* **9**, 1072–1085 (1992).
19. M. G. Lofdahl, G. B. Scharmer, and W. Wei, "Calibration of a deformable mirror and Strehl ratio measurements by use of phase diversity," *Appl. Opt.* **39**, 94–103 (2000).
20. T. J. Schulz, "Multi-frame blind deconvolution of astronomical images," *J. Opt. Soc. Am. A* **10**, 1064–1073 (1993).
21. E. J. Fernandez, L. Vabre, B. Hermann, A. Unterhuber, B. Povazay, and W. Drexler, "Adaptive optics with a magnetic deformable mirror: applications in the human eye," *Opt. Express* **14**, 8900–8917 (2006).
22. M. Born and E. Wolf, *Principles of Optics* (Pergamon, , 1986).
23. R. J. Noll, "Zernike polynomials and atmospheric turbulence," *J. Opt. Soc. Am.* **66**, 207–211 (1976).
24. Y. A. Melnikov, "Influence functions of a point force for Kirchhoff plates with rigid inclusions," *J. Mec.* **20**, 249–256 (2004).