

2019

GloPID-R report on Chikungunya, O'nyong-nyong and Mayaro virus, part I: Biological diagnostics

Michael S. Diamond
Washington University School of Medicine in St. Louis

et al

Follow this and additional works at: https://digitalcommons.wustl.edu/open_access_pubs

Recommended Citation

Diamond, Michael S. and et al, "GloPID-R report on Chikungunya, O'nyong-nyong and Mayaro virus, part I: Biological diagnostics." *Antiviral Research*. 166, 66-81. (2019).
https://digitalcommons.wustl.edu/open_access_pubs/7767

This Open Access Publication is brought to you for free and open access by Digital Commons@Becker. It has been accepted for inclusion in Open Access Publications by an authorized administrator of Digital Commons@Becker. For more information, please contact vanam@wustl.edu.



GloPID-R report on Chikungunya, O'nyong-nyong and Mayaro virus, part I: Biological diagnostics

L. Pezzi^{a,b,*}, C.B. Reusken^{c,d}, S.C. Weaver^e, J.F. Drexler^{f,g}, M. Busch^h, A.D. LaBeaudⁱ, M.S. Diamond^{j,k,l}, N. Vasilakis^m, M.A. Drebotⁿ, A.M. Siqueira^o, G.S. Ribeiro^p, A. Kohl^q, M. Lecuit^r, L.F.P. Ng^s, P. Gallian^t, X. de Lamballerie^a, on behalf of GloPID-R Chikungunya, O'nyong-nyong and Mayaro virus Working Group (S. Boyer^u, P. Brasil^v, M. Diallo^w, A.B. Failloux^x, T. Jaenisch^y, R. Lourenço-de-Oliveira^z, J. Neyts^{aa}, M. Rios^{ab}, A.J. Rodriguez-Morales^{ac}, M.G. Rosa-Freitas^z, A. Sall^w, G. Simmons^{ad}, F. Simon^{ae}, A. Vega Rúa^{af})

^a Unité des Virus Émergents (UVE: Aix-Marseille Univ-IRD 190-Inserm 1207-IHU Méditerranée Infection), Marseille, France

^b EA7310, Laboratoire de Virologie, Université de Corse-Inserm, Corte, France

^c Centre for Infectious Disease Control, National Institute for Public Health and the Environment (RIVM), Bilthoven, the Netherlands

^d Department Viroscience, Erasmus University Medical Center, Rotterdam, the Netherlands

^e Institute for Human Infections and Immunity and Department of Microbiology and Immunology, University of Texas Medical Branch, Galveston, USA

^f Charité-Universitätsmedizin Berlin, corporate member of Freie Universität Berlin, Humboldt-Universität zu Berlin, and Berlin Institute of Health, Institute of Virology, 10117 Berlin, Germany

^g German Centre for Infection Research (DZIF), Germany

^h Blood Systems Research Institute, San Francisco, and Department of Laboratory Medicine, University of California, San Francisco, USA

ⁱ Department of Pediatrics, Division of Infectious Diseases, Stanford University School of Medicine, Stanford, USA

^j Department of Medicine, and The Andrew M. and Jane M. Bursky Center for Human Immunology and Immunotherapy Programs, Washington University School of Medicine, St. Louis, USA

^k Department of Molecular Microbiology, and The Andrew M. and Jane M. Bursky Center for Human Immunology and Immunotherapy Programs, Washington University School of Medicine, St. Louis, USA

^l Department of Pathology and Immunology, and The Andrew M. and Jane M. Bursky Center for Human Immunology and Immunotherapy Programs, Washington University School of Medicine, St. Louis, USA

^m Department of Pathology, Institute of Human Infection and Immunity, University of Texas Medical Branch, Galveston, USA

ⁿ Zoonotic Diseases and Special Pathogens, National Microbiology Laboratory, Public Health Agency of Canada, Winnipeg, Canada

^o Instituto Nacional de Infectologia Evandro Chagas - Oswaldo Cruz Foundation, Rio de Janeiro, Brazil

^p Gonçalo Moniz Institute, Oswaldo Cruz Foundation, and Federal University of Bahia, Salvador, Brazil

^q MRC-University of Glasgow Centre for Virus Research, Glasgow, UK

^r Institut Pasteur, Biology of Infection Unit, Inserm U1117, Paris Descartes University, Department of Infectious Diseases and Tropical Medicine, Necker-Enfants Malades University Hospital, APHP, IHU Imagine, Paris, France

^s Singapore Immunology Network, Agency for Science, Technology and Research (A*STAR), Singapore

^t Établissement Français du Sang Alpes Méditerranée, Marseille, France

^u Medical Entomology Platform, Institut Pasteur du Cambodge, Phnom Penh, Cambodia

^v Instituto Nacional de Infectologia Evandro Chagas - Oswaldo Cruz Foundation, Rio de Janeiro, Brazil

^w Unité d'Entomologie Médicale, Institut Pasteur de Dakar, Dakar, Senegal

^x Department of Virology, Institut Pasteur, Arboviruses and Insect Vectors Unit, Paris, France

^y Section Clinical Tropical Medicine, Department of Infectious Diseases, Heidelberg University Hospital, Heidelberg, Germany

^z Instituto Oswaldo Cruz-Fiocruz, Laboratório de Mosquitos Transmissores de Hematozoários, Rio de Janeiro, Brazil

^{aa} KU Leuven, Department of Microbiology and Immunology, Rega Institute for Medical Research, Laboratory of Virology and Chemotherapy, Leuven, Belgium

^{ab} Division of Emerging and Transfusion Transmitted Diseases, Laboratory of Emerging Pathogens, Office of Blood Research and Review, Center for Biologics Evaluation and Research, U.S. Food and Drug Administration, Silver Spring, USA

^{ac} Public Health and Infection Research Group, Faculty of Health Sciences, Universidad Tecnológica de Pereira, Pereira, Colombia

^{ad} Blood Systems Research Institute, San Francisco, USA, and Department of Pathology and Laboratory Medicine, University of California, San Francisco, San Francisco, USA

^{ae} Laveran Military Teaching Hospital, Marseille, France

^{af} Laboratory of Vector Control Research, Environment and Health Unit, Institut Pasteur de la Guadeloupe, Guadeloupe

* Corresponding author. EA7310, Laboratoire de Virologie, Université de Corse-Inserm, Corte, France.

E-mail address: laura.pezzi3@studio.unibo.it (L. Pezzi).

ARTICLE INFO

Keywords:

Alphavirus
Chikungunya virus
O'nyong-nyong
Mayaro virus
Diagnostics

ABSTRACT

The GloPID-R (Global Research Collaboration for Infectious Disease Preparedness) Chikungunya (CHIKV), O'nyong-nyong (ONNV) and Mayaro virus (MAYV) Working Group is investigating the natural history, epidemiology and medical management of infection by these viruses, to identify knowledge gaps and to propose recommendations for direct future investigations and rectification measures. Here, we present the first report dedicated to diagnostic aspects of CHIKV, ONNV and MAYV. Regarding diagnosis of the disease at the acute phase, molecular assays previously described for the three viruses require further evaluation, standardized protocols and the availability of international standards representing the genetic diversity of the viruses. Detection of specific IgM would benefit from further investigations to clarify the extent of cross-reactivity among the three viruses, the sensitivity of the assays, and the possible interfering role of cryoglobulinaemia. Implementation of reference panels and external quality assessments for both molecular and serological assays is necessary. Regarding sero-epidemiological studies, there is no reported high-throughput assay that can distinguish among these different viruses in areas of potential co-circulation. New specific tools and/or improved standardized protocols are needed to enable large-scale epidemiological studies of public health relevance to be performed. Considering the high risk of future CHIKV, MAYV and ONNV outbreaks, the Working Group recommends that a major investigation should be initiated to fill the existing diagnostic gaps.

1. Introduction

The GloPID-R (Global Research Collaboration for Infectious Disease Preparedness) Chikungunya (CHIKV), O'nyong-nyong (ONNV) and Mayaro virus (MAYV) Working Group presents a report dedicated to diagnostic aspects of these pathogens. Overlapping clinical presentations, wide geographic spread and the possible selection of viruses capable of transmission by new vectors highlight the need for laboratory diagnostic support to unambiguously identify etiological agents in cases of undefined febrile illness with arthralgia and/or rash. This is important to ensure an early detection of cases and to support a clinical and public health response.

In order to assess the currently available molecular and serological tools to diagnose infections by CHIKV, ONNV and MAYV, the experts of the GloPID-R have performed a systematic review of English literature on the diagnostic aspects of the three viruses present on PubMed until September 2018. Diagnosis of the diseases at the acute phase mostly relies on molecular detection of the virus genomes: we discussed about kinetics of viral loads and biological sampling, choice of molecular tool, available in-house and commercial molecular tests, international standards and external quality assessments organized to investigate laboratories' capability for viral detection. Alternatively, diagnosis can be provided by the detection of specific antibodies, so we focused on kinetics of immune response, choice of serological tool, commercially available tests, international standards and external quality assessments; moreover, cryoglobulinaemia and cross-reactivity, that can interfere with the correct identification of antibodies, were discussed, as well as cross-protection and cross-neutralization among the three pathogens.

Viral isolation by culture has not been taken into account because its turnaround time hardly fits with the need for an early diagnosis. Although in some cases (*i.e.* 2005–2006 CHIKV outbreak in Mauritius [Pyndiah et al. \(2012\)](#)) virus isolation has been performed for laboratory diagnosis with good sensitivity, providing results 2–3 days after inoculation, this diagnostic modality is being replaced by molecular techniques including RT-PCR, that are faster, more sensitive and safer and require less laboratory infrastructure. However, virus isolation remains a crucial tool in reference laboratories to characterize circulating strains to support control and development of diagnostic tools.

In this report, the experts identified knowledge gaps and provided adapted recommendations. The objective was to suggest relevant research priorities in the field, in order to improve individual patient care and outbreak management.

2. Background

2.1. Epidemiology

At least 5 million cases over the last 15 years ([Chikungunya, 2018](#)) are reason enough to recognize *Chikungunya virus* (CHIKV) as an emerging global health threat. Data from the two largest CHIKV epidemics so far highlight the ability of the virus to spread rapidly over great distances and in multiple locations. During the 2004–2006 outbreak in the Indian Ocean region, attack rates peaked at 63% in the Comoro Islands; La Réunion Island registered 47,000 cases in a single week (January 30th to February 5th, 2006), with almost 40% of the population estimated to have experienced CHIKV disease; India recorded at least 1.4 million cases at the end of 2006. In late 2013, CHIKV emerged in the New World, when in the island of Saint Martin CHIKV infection was identified in patients with no history of travel abroad. The virus spread to 45 countries and territories in North, Central, and South America and the Caribbean Islands; by the end of 2017, 544 deaths directly or indirectly related to the outbreak in the Americas were recorded, together with 2.5 million cases of infections. Transmission is ongoing in some locations ([PAHO WHO, 2019](#)).

CHIKV is a mosquito-transmitted alphavirus belonging to the family *Togaviridae* ([Powers and Logue, 2007](#)). The two major vectors of the disease are *Aedes (Stegomyia) aegypti* and *Ae. (Steg.) albopictus* ([Vega-Rua et al., 2014](#)), the latter identified during the 2004–2006 outbreak in the Indian Ocean, when the first in a series of adaptive mutations in the glycoprotein genes of the virus increased its replication in this specific vector ([Schuffenecker et al., 2006](#); [Tssetsarkin et al., 2016](#)). Two transmission cycles have been described. A sylvatic cycle maintains CHIKV in Africa involving forest-dwelling mosquito species (such as *Ae. africanus* and *Ae. furcifer-taylori*) and mainly nonhuman primates ([Althouse et al., 2018](#)); in this context, humans are incidental hosts and become infected when they enter the forest or when infected vectors invade adjacent villages from the forest. Several wild animal species, especially nonhuman primates, have been investigated as possible virus reservoirs sustaining virus circulation in the environment in the absence of human cases ([Diallo et al., 1999](#)). CHIKV can also be maintained through an urban human-mosquito-human transmission cycle, that involves *Ae. aegypti* and *Ae. Albopictus* mosquitoes; this cycle has been observed in the Americas, Indian Ocean, Asia and, more recently, Europe. In this urban context, humans can serve as reservoir and amplification hosts ([Thiberville et al., 2013a](#)).

Several factors contribute to massive CHIKV circulation: *e.g.* high mosquito densities, including following recent invasions of the main

urban vectors, *Aedes* species; vectorial capacity in transmitting the virus (high susceptibility to CHIKV as well as preference for feeding on human beings); high viremia level in infected humans; large population of susceptible individuals exposed for the first time in several countries; increasing travel from countries with ongoing circulation that enables importation of viremic cases into virus-naïve regions, allowing the virus to extend its geographic distribution if conditions for local transmission are present.

CHIKV forms, with seven other alphaviruses, the Semliki Forest serocomplex, in which E1 envelope glycoprotein gene amino acid sequence divergence is below 40% (Powers et al., 2001). Among this serocomplex, viruses from the species *Mayaro virus* (MAYV) and *O'nyong-nyong virus* (ONNV) are of particular interest because they have potential to emerge in the human population (Hotez and Murray, 2017; Lwande et al., 2015).

ONNV is transmitted by anopheline mosquitoes; it is probably maintained through an enzootic cycle that has not been characterized yet, and humans can serve as amplification hosts during epidemics (Rezza et al., 2017). MAYV is transmitted by *Haemagogus* species mosquitoes in an enzootic cycle in which nonhuman primates seem to be the main amplifier vertebrate hosts, with limited spillover to humans who frequently enter forest habitats (Mackay and Arden, 2016).

CHIKV, ONNV and MAYV are phylogenetically related: they share some features that make differential diagnosis among the three viruses a challenging task.

2.2. Clinical presentation

Infections by CHIKV, ONNV and MAYV share an apparent similar clinical picture, with a “dengue-like” syndrome (Thiberville et al., 2013a; Rezza et al., 2017; Mackay and Arden, 2016). This typically includes fever, headache and a papular or maculopapular rash during the acute stage, together with a more specifically remarkable incapacitating polyarthralgia, which is the hallmark of these so-called “alphavirus arthritogenic diseases”. Although disease is generally self-limiting, severe joint pain can persist for months or even years in some individuals (Heath et al., 2018).

2.3. Co-circulation

CHIKV impacts human health globally, with local transmission reported in Asia, Africa, Europa, Americas, Pacific region and, transiently, in Europe (Thiberville et al., 2013a; Lwande et al., 2015). MAYV circulation has mostly been limited to South America with sporadic outbreaks next to forest environments (Schmidt et al., 1959; Anderson et al., 1957; LeDuc et al., 1981). Recently detected MAYV infections in Haiti in a patient with no history of travel abroad living in a non-forest area, and in Panama suggest (I) the spread of the virus towards Central America and the Caribbean and (II) a possible human-to-human transmission (Lednický et al., 2016; Carrera et al., 2018). ONNV is endemic in sub-Saharan Africa and it has been associated with sporadic large-scale epidemics. The first one, in 1959–1962, accounted for more than 2 million cases in eastern Africa alone (Haddow et al., 1960); then, after an apparent absence of about 35 years, ONNV re-emerged in Uganda in 1996 (Rwaguma et al., 1997), showing again its potential to cause massive outbreaks. Moreover, the actual distribution of ONNV in Africa is likely to be broader. Limited availability of diagnostic tools may have led to confusion of ONNV and CHIKV outbreaks (LaBeaud et al., 2015).

Co-circulation of CHIKV-MAYV in Southern and Central America and CHIKV-ONNV in the African continent, do not allow to make a reliable diagnosis based on the geographic area where the infection is contracted; moreover, the wide circulation of other pathogens causing febrile illness (such as dengue virus and malaria-causing *Plasmodium* species) makes a travel history of the patient insufficient to narrow the differential diagnosis (Cleton et al., 2012).

2.4. Transmission vectors

The three viruses do not share the same main mosquito vectors: CHIKV is primarily transmitted by *Aedes* spp; ONNV is unique among alphaviruses in its adaptation to *Anopheles* spp; MAYV enzootic vectors are mostly *Haemagogus* spp (Coffey et al., 2014; Williams et al., 1965; Hoch et al., 1981). However, experimental evidence that *Aedes* mosquitoes can transmit ONNV and MAYV exists (Vanlandingham et al., 2005; Long et al., 2011; Smith and Francy, 1991), with a productive viral replication both *in vitro* and *in vivo*. As CHIKV has taught us, a single amino-acid change in the virus can result in increased vector competence from a mosquito species not normally considered a primary vector (Tssetsarkin et al., 2007). The adaptation of CHIKV leading to higher competence of *Ae. albopictus* for virus transmission among humans illustrates that virus evolution might drive the involvement of other, more anthropophilic mosquito species.

3. Chikungunya virus (CHIKV)

3.1. Molecular diagnosis

3.1.1. Kinetics of viral load and dedicated biological sampling

CHIKV infections are characterized by high viremia levels (range: 10^4 – 10^8 RNA copies/mL (Laurent et al., 2007; Lanciotti et al., 2006)) during the acute phase, which includes the first 3–4 days after illness onset (Chow et al., 2011; Thiberville et al., 2013b). This stage is followed by a rapid decrease of viremia (Chusri et al., 2014): in 2008, Panning et al. (2008) showed that real-time RT-PCR was 100% positive for all sera collected up to day 4 after illness onset, with sensitivity subsequently decreasing to 40% at day 7. Similarly, CHIKV loads have been proved to decrease significantly in serum from 5 days after the onset of symptoms onwards (Bozza et al., 2019). Even if some persisting cases of viremia have been observed (up to 17 days) (Win et al., 2010; Appassakij et al., 2013; Leo et al., 2009), molecular diagnosis should typically be performed within the first 7 days on an acute-phase specimen to confirm CHIKV infection (Appassakij et al., 2013; Reddy et al., 2012; Musso et al., 2016).

The preferred samples for diagnosis are serum or plasma (Lanciotti et al., 2006; Panning et al., 2008; Appassakij et al., 2013; Musso et al., 2016; Diagnostic Testing, 2017), but other clinical samples have been evaluated for use in diagnosis.

3.1.1.1. Saliva. Saliva showed a lower sensitivity compared to blood for CHIKV: it has been observed that for confirmed CHIKV cases (by IgM or PCR in blood (WER9033_410-414)) presenting during the 1st week after illness onset, detection rate in saliva was 58,3% compared to 86,1% in blood samples (Musso et al., 2016). Accordingly, testing saliva can be relevant during the 1st week post-symptoms onset if blood samples are difficult or impossible to collect, but is associated with decreased detection rate and lower negative predictive value.

3.1.1.2. Semen. Interestingly, semen has been found positive for a prolonged period of time after symptom onset (30 days), providing an expanded window for the molecular diagnosis of a recent infection with CHIKV (Bandeira et al., 2016). This finding needs to be further investigated, in order to assess the risk of sexual transmission of CHIKV and how this risk compares to the recently observed sexual transmission of other mosquito-borne viruses such as Zika virus (Mead et al., 2018).

3.1.1.3. Urine. The same study showed CHIKV RNA 30 days after initial symptoms in urine from a patient with a positive semen sample (Bandeira et al., 2016). Similarly, viral genomes were present up to day 30 post-infection in the urine of infected mice, long after viral clearance from the plasma (Jones and Okeoma, 2015). Recently, a case report described a patient with meningoencephalitis and ocular lesions having a positive urine sample 40 days after illness onset (Rocha et al.,

2018). However, Musso et al. (2016) did not observe a prolonged window of detection of CHIKV in urine, with samples testing positive by PCR only if collected during the first week after symptoms onset. For confirmed CHIKV patients presenting during the first 7 days after illness onset, the detection rate in the urine was 8.3% compared to 86.1% in blood; after the 1st week after symptom onset, it decreased to 0%, with no positive urine samples that could confirm serological diagnosis of CHIKV infection (IgM+). Bozza et al. observed low viral loads in the urine of patients with confirmed CHIKV infection (Bozza et al., 2019). Considering the contradictory results, virus shedding kinetics in urine deserve a systematic re-assessment with enough power in the cohort size.

3.1.1.4. Breast milk. Recently, the presence of CHIKV RNA in breast milk was reported 23 days after the inception of symptoms in one patient (Campos et al., 2017); however, another study could not detect viral RNA in maternal milk in 8 samples collected during viremia (although plasma samples from the same patients were RT-PCR positive) (Grivard et al., 2007).

3.1.1.5. Cerebrospinal fluid. In cases of CHIKV infections with neurological involvement, cerebrospinal fluid (CSF) samples can be tested with both serological and virological assays, with several studies describing detection of viral RNA in CSF (Rocha et al., 2018; Grivard et al., 2007; Gérardin et al., 2016; Lemant et al., 2008; Chandak et al., 2009; Nagpal et al., 2017; Taraphdar et al., 2015). Case reports suggest that genomes can persist in the CSF for 7–10 days after illness onset (Taraphdar et al., 2015; Casolari et al., 2008), with viral loads in the CSF generally lower than in plasma (Grivard et al., 2007).

3.1.1.6. Other fluids and tissues. The virus has been detected also in amniotic fluid, brain and liver biopsies, neonatal gastric fluids and placenta (Grivard et al., 2007). Windows of detection of CHIKV genomes in different samples are presented in Fig. 1.

3.1.2. Choice of test

Several diagnostic methods are available for viral genome detection. Molecular assays targeting nucleic acids are sensitive methods in the early stages of infection, before the rise of antiviral IgM antibodies (Lanciotti et al., 2006). Isothermal amplification methods have been described for CHIKV, *i.e.* loop-mediated isothermal amplification (LAMP) and transcription-mediated amplification (TMA) (Lu et al., 2012; Parida et al., 2007; Chiu et al., 2015). However, real-time RT-PCR remains the most frequently used technique for routine diagnosis at the acute stage of infection, because of its sensitivity, specificity and ease of performance (Reddy et al., 2012). A positive result by real-time RT-PCR is enough to make a reliable diagnosis of CHIKV, ONNV or MAYV in areas where virus circulation has been previously documented. However, in case the virus is detected for the first time in a new geographic area, it is good practice to confirm a positive result by RT-PCR with a different test, to rule out a false-positive result caused by laboratory contamination. Conventional RT-PCR is still employed for research purposes (*i.e.* to identify to which lineage a viral strain belongs (Lednicky et al., 2016; Bessaud et al., 2006; Calba et al., 2017)) but is considered to be outdated for routine diagnostics.

3.1.3. In-house real-time RT-PCR

Several real-time RT-PCR systems have been published (Lanciotti et al., 2006; Panning et al., 2008; Waggoner et al., 2016; Cecilia et al., 2015; Smith et al., 2009; Edwards et al., 2007; Pastorino et al., 2005). Some of them have been evaluated in the External Quality Assessment (EQA) organized in 2014 by Jacobsen et al. (2016); their performances are discussed in the paragraph 3.1.6. An alignment of 50 CHIKV sequences available in Genbank representing main lineages was made to assess published in-house developed primer sets; GenBank accession numbers and results of this *in silico* analysis are presented in Table 1.

Most PCR systems show one or more mismatches with CHIKV strains in different proportion and positions; particular attention was given to mismatches concerning the five 3' terminal nucleotides of a primer, because it is generally admitted that a single mutation in these positions can significantly compromise detection capability of a primer. The table highlights the fact that a number of available PCR systems are expected to detect viruses from the WA lineage less efficiently than viruses from the other lineages.

3.1.4. Commercial molecular tests

Several commercial tests from different companies were identified for CHIKV molecular detection by real-time RT-PCR (Supplementary data-Table 1). Some of them are formulated in monoplex format, whereas others are multiplex assays targeting dengue virus (DENV) and Zika virus (ZIKV) as well as in some cases yellow fever virus (YFV) and usutu virus (USUV). RealStar kit from Altona Diagnostics (Hamburg, Germany) is the only ready-to-use kit for real time RT-PCR evaluated in the literature through the comparison with a published real-time RT-PCR protocol (Panning et al., 2008, 2009a). The target region is within the non-structural protein 1 (nsP1) gene and the manufacturer claims that the kit allows detection of all three CHIKV genotypes; however, lack of data on primers and probe sequences does not allow for an *in silico* analysis to confirm this information, and only the ECSA genotype was tested in the evaluation study. The 95% limit of detection (LOD) of the kit is 3.2 genome copies per reaction using a quantified RNA from CHIKV ECSA strain (RealStar, 2018). The evaluation study defined the LOD using two different thermocyclers, Lightcycler 2.0 and Lightcycler 480. Using *in vitro*-transcribed RNA copies of a fragment of an ECSA strain, 95% LOD was 5.3 copies per reaction with Lightcycler 2.0 and 3.8 copies per reaction with Lightcycler 480. Using a plaque-purified and plaque-quantified CHIKV ECSA strain, LOD was 0.51 PFU/mL with Lightcycler 2.0 and 0.34 PFU/mL with Lightcycler 480. High specificity was established by the absence of cross-reactivity against a large panel of non-CHIKV alphaviruses (including MAYV and ONNV) and non-alphaviruses. The RealStar assay had 100% sensitivity and specificity when compared with the previously published real-time RT-PCR of Panning et al. (2008), which has shown good sensitivity (95%) and specificity (87,5%) values in the context of a multi-partner External Quality Assessment (Jacobsen et al., 2016).

The RealStar kit has been used for testing blood donations prior to transfusion as well as for CHIKV diagnosis in patients. The French Blood Agency used it during the 2014 CHIKV Caribbean outbreak for




Fig. 1. Windows of detection of CHIKV genomes in different kinds of samples by molecular methods. The letters in brackets indicate the references. (a): (Lanciotti et al., 2006; Panning et al., 2008; Musso et al., 2016); (b): (Win et al., 2010; Appassakij et al., 2013); (c): (Appassakij et al., 2013); (d): (Musso et al., 2016); (e): (Bandeira et al., 2016; Jones and Okeoma, 2015); (f): (Rocha et al., 2018); (g): (Musso et al., 2016); (h): (Bandeira et al., 2016); (i): (Casolari et al., 2008); (j): (Taraphdar et al., 2015); (k): (Campos et al., 2017).

Table 1

Ten sequences were selected for each lineage (CHIKV WA, GenBank accession numbers: HM045815, HM045818, AY726732, HM045785, HM045798, HM045786, HM045807, HM045819, HM045820; CHIKV ECSA: AF369024, AF490259, KJ679577, KP164570, KY704947, HM045822, HM045806, HM045795, HM045792, JQ067624; CHIKV IOL: GQ428211, FJ807896, FJ000062, FJ807899, GQ428212, EF027136, EU564334, EF012359, HQ456254, MG664850; CHIKV Asian FJ807897, HM045791, HM045787, HM045810, HM045788, EF027140, EU703759, KT308159; CHIKV Asian/American: KR559497, KR559496, KR559473, KR559492, KJ451624, KR046227, KR046231, LN898098, LN898093, KR559493). Mismatch positions between this reference sequence panel and the primers and probes are indicated in the table; mismatches concerning the five 3' terminal nucleotides of a primer are indicated in red and bold characters. In probes, 1 or 2 mutations generally do not compromise the hybridization potential, but this also depends on the length of the probe and its G + C content.

In house real-time PCR systems		CHIKV WA (nb seq = 10)	CHIKV ECSA (nb seq = 10)	CHIKV IOL (nb seq = 10)	CHIKV ASIAN (nb seq = 10)	CHIKV ASIAN/AMERICAN (nb seq = 10)
Target(s)	Reference	Technique	Amplicon size (bp)	FW primer 5'-3'	RV primer 5'-3'	Probe
NSP2	Waggoner et al. (2016)	Hydrolysis probe	96	CATCTGCACYCAAGTGTACCA	GCGCATTTTGGCCTTGGTAATG	GCGGTGTACACTGCCTGTGACYGC
E3	Cecilia et al. (2015)	Hydrolysis probe	64	CGAAAARGARCGGAGRAA	GATAGTACCCRRGKCTCATGACGTT	CGAAAARGARCGGAGRAA
E1	Smith et al. (2009)	Hydrolysis probe	253	CCCTRGGCATGCTTGA	CCACATAAATGGGTAGACGCC	CCCTRGGCATGCTTGA
NSP1	Smith et al. (2009)	Hydrolysis probe	74	TTTGTGATCAAAITGACGGGCATC	TCAGCCCGCACCAACAGCTTC	TTTGTGATCAAAITGACGGGCATC
NSP2	Smith et al. (2009)	Hydrolysis probe	70	CCGAAAGGAACTTCAAAGCAACT	TTGCTACAGAAAGTAC	CCGAAAGGAACTTCAAAGCAACT
5'UTR	Smith et al. (2009)	Hydrolysis probe	98	ACACAGGTAGCCTACCAGTTTC	GCTGTACAGCGTCTATGTCCAC	ACACAGGTAGCCTACCAGTTTC

Mismatch positions in primers and probe; the 3' terminal position corresponds to position 1; mismatches in the five 3' terminal nucleotides are indicated in red and bold; (nb of CHIKV sequences with mismatch in the specific lineage considered)

Pos 9: A > G (1)
 Pos 7: G > A (2)
 Pos 14: A > G (2)
 Pos 19: C > T (2)
 Pos 1: A > C (10)
 Pos 2: A > G (10)
 Pos 4: G > A (10)
 Pos 6: G > C (10)
 Pos 18: C > T (10)
 Pos 24: A > G (10)
 C > T (10)
 Pos 3: G > T (10)
 C > T (10)
 C > T (3)
 Pos 9: G > A (1)
 Pos 15: A > G (10)
 C > T (3)
 T > C (1)
 Pos 13: T > C (1)
 Pos 4: C > T (1)
 Pos 3: G > A (3)
 C > T (10)
 Pos 15: A > G (10)
 C > T (10)
 T > C (1)
 Pos 7: A > G (7)
 Pos 9: A > G (10)
 Pos 5: T > C (10)
 Pos 21: A > G (1)
 T > C (10)
 Pos 7: A > G (10)
 Pos 5: T > C (10)
 Pos 21: A > G (1)
 T > C (10)
 Pos 8: C > T (1)
 Pos 16: G > A (1)
 Pos 5: G > T (1)
 Pos 6: A > G (1)
 Pos 11: C > T (1)
 Pos 16: G > A (1)
 Pos 8: C > T (1)
 Pos 6: G > A (7)
 Pos 15: A > G (10)
 - > A (1)

(continued on next page)

Table 1 (continued)

In house real-time PCR systems		CHIKV WA (nb seq=10)	CHIKV ECSA (nb seq=10)	CHIKV IOL (nb seq=10)	CHIKV ASIAN (nb seq=10)	CHIKV ASIAN/AMERICAN (nb seq=10)
NSP1	Panning et al. (2008)	FW primer 5'-3' TGATCCGCACTCAACCAATCCT	Pos 9: G > A (2)	Pos 18: T > C (10)		
NSP1	Lanciotti et al. (2006)	RV primer 5'-3' GGCAAAACGAGTGTACTTCTCT	Pos 18: A > G (4)	Pos 18: A > G (4)		
		Probe 5'-3' GCCAGCAAGGAGGATGATGTCGGA AAAGGGCAAACTCAGCTTCAC	G > A (2) Pos 12: A > G (10) Pos 15: C > G (10) Pos 18: G > A (10) Pos 11: C > A (10) Pos 17: T > C (10)		Pos 9: C > T (9) Pos 12: A > G (9)	G > A (1) Pos 9: C > T (9) Pos 12: A > G (9)
NSP4	Lanciotti et al. (2006)	RV primer 5'-3' GCCTGGGCTCATCGTTATTC	Pos 13: C > T (1) Pos 14: G > A (1) T > G (10)		Pos 2: T > C (1) Pos 17: T > C (1) T > C (9)	
		Probe 5'-3' CGCTGTGATACAGTGGTTTCGTGTG TCACTCCCTGTGGACTTGAAGA TTGAGGAAACAGAGTTAGGAACATACC	Pos 20: T > C (10) Pos 21: C > T (10) Pos 9: G > A (3) Pos 10: G > A (1) Pos 15: A > G (10) A > G (10) T > C (10) G > T (1) C > T (10) C > T (10)	Pos 14: T > C (10)	Pos 8: T > C (3) Pos 17: C > T (2)	Pos 8: T > C (10) Pos 17: C > T (10)
E1	Edwards et al. (2007)	Probe 5'-3' AGGTACGGGCTTCAAGTTCGGGG	C > T (1)		C > T (1)	
		FW primer 5'-3' TCGAGCGCCCTCTTTAA	Pos 4: T > G (10) Pos 11: G > A (10) Pos 14: C > T (10)		Pos 5: T > C (1) Pos 14: C > A (1)	Pos 8: C > A (10)
E1	Pastorino et al. (2005)	RV primer 5'-3' ATCGAATGCACGGCAGACT				
		Probe 5'-3' ACCAGCCTGCACCCGATTCCTCAGAC	C > T (10) T > C (1) A > C (10)		C > T (2)	C > T (10) A > G (1)
E1	Pastorino et al. (2005)	FW primer 5'-3' AAGCTYCGGTCCTTTAGCAAAG	Pos 5: C > T (1)			
		RV primer 5'-3' CCAAAATTGTCYGGTCTTCTCT				Pos 16: T > C (1)
		Probe 5'-3' CCAATGCTCTMGGCTGGACACCTTT	C > G (1)			

Mismatch positions in primers and probe; the 3' terminal position corresponds to position 1; mismatches in the five 3' terminal nucleotides are indicated in red and bold; (nb of CHIKV sequences with mismatch in the specific lineage considered)

individual NAT (nucleic acid testing) screening of blood products as a part of implemented strategy to prevent CHIKV transfusion-transmitted infections (TTIs) (Gallian et al., 2014).

According to the datasheet of the company, the FTD DENV/CHIKV real time PCR kit from Fast-track Diagnostics has been evaluated with EQA panels from Quality Control for Molecular Diagnostics (QCMD) (Dengue Fever PCR Test, 2018); all samples (LeDuc et al., 1981) were detected correctly except one.

During the 2014 CHIKV outbreak in the Americas, a prototype of real-time CHIKV/DENV target-capture, transcription-mediated amplification (TC-TMA) assay by Hologic, Inc. (San Diego, CA, USA) was used for high-throughput screening of blood products (Simmons et al., 2016) on the Panther system (Hologic, Inc.). Tests were conducted during and after the CHIKV epidemic in Puerto Rico in 2014 to detect Asian genotype viruses in plasma samples. The Panther system automates all aspects of NAT-based blood screening on a single, integrated platform; it is based on Procleix technology (developed by Hologic and Grifols), already adapted for transfusion screening for HIV, HBV and HCV. During the epidemic, high sensitivity was evidenced for both individually tested samples and 16-sample minipools, with a 95% LOD of 9.9 and 158 copies per reaction, respectively.

The same TC-TMA assay was used by another group in the same period (2014 CHIKV Caribbean epidemic) (Chiu et al., 2015), with a LOD of 8.2 copies/reaction. Positive TMA results were confirmed by PCR, microarray and next-generation sequencing (NGS) analysis.

3.1.5. International standard (IS)

The first CHIKV RNA World Health Organization (WHO) International Standard (IS) for Nucleic Acid Amplification Techniques (NAAT)-Based Assays was accepted in October 2017 and is distributed by the Paul Ehrlich Institute (Germany) (PEI, 2018). It has been prepared from strain R91064 of the East/South/Central African (ECSA) genotype (Indian Ocean lineage), isolated from a patient returning from India to the USA in 2006 (Lancioti et al., 2006). The accession number for the Genbank sequence is KJ941050 (Añez et al., 2014). The freeze-dried preparation contains heat-inactivated virus that has been diluted in human plasma negative for anti-CHIKV antibodies. This reagent has been assigned a unitage of 2,500,000 International Units/mL, when reconstituted as recommended in 0.5 mL of sterile nuclease-free water. The material has been evaluated in an international collaborative study involving 25 laboratories performing a wide range of CHIKV NAAT assays (who-collaborating-study-chikungunya).

A CHIKV RNA reference reagent (RR) was produced in 2015 by the CBER/FDA (Center for Biologics for Evaluation and Research/US Food and Drug Administration) (Añez et al., 2015a). It is now available from the Office of Blood Research and Review, CBER/FDA (Research C forand. Bl, 2018). The reference reagent consists of cell culture-grown, heat-inactivated CHIKV diluted in human plasma and frozen. The strain is also the above-mentioned R91064. In a collaborative study involving 8 laboratories, the RR showed an estimated overall mean of 7.56 log₁₀ detectable units/mL, ranging from 6.2 log₁₀ to 8.6 log₁₀ (Añez et al., 2015b).

International Standards based on West African and Asian lineages have not been developed. This is of specific importance, because Asian lineage of CHIKV has been broadly circulating in Asia, Latin America and the Caribbean, and there are still episodes of circulation of the West African lineage of CHIKV in Senegal and most probably in neighboring countries (Sow et al., 2017; Eastwood et al., 2017).

3.1.6. External quality assessment (EQA)

Four EQAs (Table 2) have been organized to investigate laboratories' capability for detection of CHIKV RNA by molecular tools, three of which were coordinated by the European Network for Diagnostics of Imported Viral Diseases (ENIVD) (now EVD-LabNet).

Donoso-Mantke et al. (Donoso Mantke and Niedrig, 2007) organized in 2007 a study involving 24 European laboratories to assess both

Table 2
EQAs for molecular detection of CHIKV RNA.

Year	Reference	Participants	Study coordinator	Objectives	Other viruses tested in the panels	Results of molecular assessment
2007	Donoso Mantke and Niedrig (2007)	Total: 24 labs from 15 European countries//20 labs for CHIKV PCR	European Network for Diagnostics of Imported Viral Diseases (ENIVD)	Assessing quality for CHIKV molecular and serological diagnostics	-	16/20 labs showed sufficient sensitivity and specificity
2007	Panning et al. (2009b)	Total: 31 labs from Europe, Asia, Africa, South America	European Network for Diagnostics of Imported Viral Diseases (ENIVD)	Coordinated Implementation of CHIKV RT-PCR	-	14/31 labs met proficiency criteria
2014	Jacobsen et al. (2016)	Total: 56 labs from 40 countries (Europe, Asia, Africa, Middle East, Americas, Caribbean, Oceania)	European Network for Diagnostics of Imported Viral Diseases (ENIVD)	Assessing quality for CHIKV molecular and serological diagnostics	SINV, ONNV, DENV	23/60 data sets classified as "optimal", 7/60 "acceptable", 30/60 "need for improvement"
2015	Soh et al. (2016)	Total: 24 labs from 22 countries and areas (South-East Asia, Western Pacific Regions)//20 labs for CHIKV PCR	-	Assessing quality for CHIKV molecular diagnostic and for DENV molecular and serological diagnostics	DENV	19/20 labs reached the highest score

molecular and serological diagnostics for CHIKV. Testing a panel by molecular tools, a great variability in performances was observed; in particular, 4 of the 20 laboratories that participated used assays with low sensitivity and specificity.

In 2007, Panning et al. (2009b) organized an international proficiency study with 31 participants from Europe, Asia, Africa and South America, who were asked to test the material with any molecular assay routinely used for detecting CHIKV in human plasma or with a real-time PCR protocol previously distributed by the study coordinator. In total, 17/31 laboratories lacked adequate sensitivity; those who used the assay distributed by the consortium showed the highest performances in term of sensitivity, proving that *ad hoc* improvement of molecular diagnostics was possible. False-positive results obtained with nested RT-PCR, a technique often affected by risk of contamination, confirmed it as obsolete for routine diagnosis.

In 2014, a third EQA (by Jacobsen et al.) involved 56 international laboratories from Europe, Asia, Africa, Middle East, Americas, Caribbean, Oceania (Jacobsen et al., 2016), allowing for worldwide performance evaluation. Conventional in-house RT-PCR tests had the highest rate of correct results, although only 5 laboratories used it; however, low detection rate was observed for the lowest viral load of the dilution series. In-house real-time RT-PCR systems were the most commonly used techniques; in particular, systems from Pastorino (Pastorino et al., 2005) and Panning (Panning et al., 2008) were used by 9 and 8 out of 42 labs, respectively. As other PCR systems used in the EQAs (by Lanciotti and Edwards, both used by 3 labs (Lanciotti et al., 2006; Edwards et al., 2007)), they showed heterogeneous results with laboratories using the same protocol especially in terms of sensitivity, suggesting problems associated with laboratory procedures; specificity and genotyping were more often correctly evaluated. Some commercial real-time RT-PCRs (used by 13 labs) raised issues for their lack of sensitivity and specificity (with the closely related ONNV frequently found positive); however, the identification of these poorly performing kits was not provided.

The most recent EQA was conducted in 2015 by Soh et al. (2016) among 24 national-level public health laboratories in the Asia Pacific region to assess both CHIKV and DENV diagnostics. The majority of them requested receipt of a CHIKV-positive control and/or real-time or conventional RT-PCR protocols to develop and validate their capacity for CHIKV diagnosis. High degrees of sensitivity and specificity were observed, with 19/20 laboratories (95%) having detected correctly CHIKV; the only one that detected CHIKV in a serum-only plasma used a real time RT-PCR protocol.

3.2. Serological diagnosis

3.2.1. Kinetics of the immune response

Virus-specific IgM antibodies appear in the serum within 4–6 days after the onset of illness (Chusri et al., 2014; Staples et al., 2009). Accordingly, IgM may be absent at the initial consultation, at the acute stage of the disease. They generally can be detected up to 3–4 months after infection (Chusri et al., 2014; Staples et al., 2009; Prince et al., 2015), but can persist for more than one year, especially for patients with chronic arthralgia (Malvy et al., 2009; Borgherini et al., 2007). In case of CHIKV infections with neurological involvement, cerebrospinal fluid (CSF) can be tested for IgM antibodies (Tournebize et al., 2009; Mehta et al., 2018). According to case reports, they would appear a little later than in serum (Taraphdar et al., 2015); moreover, the higher the IgM titer in the serum, the higher it is in CSF (Gérardin et al., 2016).

IgG antibodies are typically found as early as 6–7 days after illness onset, a few days after IgM appearance (Chusri et al., 2014; Prince et al., 2015); like some IgM antibodies, they can directly neutralize CHIKV multiplication and can persist in immune individuals for many years (Nitapattana et al., 2014). In the case of a strong and long-lasting IgM response, it seems that IgG seroconversion can occur late, with no IgG detected in some patients 90 days after symptom onset (Bozza et al., 2019). Among all four IgG isotypes, IgG3 antibodies dominate in the naturally-acquired IgG response, and they are mostly specific for the E2 protein (Kam et al., 2012a; Verma et al., 2013). Their appearance during the early convalescent phase has been associated with virus clearance, long-term clinical protection and better outcome (Kam et al., 2012b).

Cross-reactivity of CHIKV antibodies has been observed with members of Semliki Forest serocomplex, especially with viruses having a close genetic and evolutionary relationship (ONNV, MAYV, ross river virus-RRV) (Calisher et al., 1986; Karabatsos, 1975; Strauss and Strauss, 1994). It will be discussed more in detail in the paragraph 3.2.8.

3.2.2. Choice of test

After the period of viremia has ended, diagnosis may rely on virus detection in other fluids or tissues (see above) or on serological assays. A variety of laboratory diagnostic methods have been developed over time, including immunofluorescence tests (IFT), haemagglutination-inhibition assays (HIA), enzyme-linked immunosorbent assays (ELISA) and neutralization tests (NT).

ELISAs are the most commonly used assays for detection of both IgG and IgM antibodies. ELISA and IFT, in contrast to HIA and NT, can make the distinction between IgM and IgG; this is useful because the detection of IgM, or detection of IgG seroconversion or a four-fold rise

Table 3

Performances of commercial tests for the detection of anti-CHIKV IgM antibodies, according to published evaluations.

Kit	Company	Method	Performance (ref)
Human Anti-CHIKV IgM ELISA Kit	Abcam	IgM ELISA	Lot-to-lot variation: good concordance with CDC results vs low sensitivity (Johnson et al., 2016)
CHIK IgM ELISA Test CE	CTK Biotech	IgM ELISA	CDC: Low sensitivity (Johnson et al., 2016)
Chikungunya IgM Combo Rapid Test CE		IgM-rapid test	CDC: Low sensitivity (Johnson et al., 2016; Prat et al., 2014; Kosasih et al., 2012; Yap et al., 2010)
Anti CHIKV IFT	EUROIMMUN	IFT	CDC: High accuracy and reproducibility; Good sensitivity and specificity; Variation in sensitivity for Ab against different CHIKV strains (Johnson et al., 2016; Yap et al., 2010; Litzba et al., 2008)
Anti-CHIKV ELISA (IgM)		IgM ELISA	CDC: High accuracy and reproducibility; Cross-reactivity with anti-ONNV Abs; High sensitivity; False neg and false pos results (Johnson et al., 2016; Prat et al., 2014; Prince et al., 2016)
Chikungunya IgM μ -capture ELISA	GenWay	IgM ELISA	CDC: Low sensitivity (Johnson et al., 2016)
Chikungunya IgM μ -capture ELISA	IBL International	IgM ELISA	Cross reactivity with anti-ONNV Abs; False neg and false pos results (Prat et al., 2014)
CHIKij Detect™ IgM ELISA Kit	InBios	IgM ELISA	CDC: High accuracy and reproducibility; High sensitivity (Johnson et al., 2016; Prince et al., 2016)
SD Chikungunya IgM ELISA	SD Standard Diagnostics	IgM ELISA	CDC: Low sensitivity (Johnson et al., 2016; Blacksell et al., 2011)
SD BIOLINE Chikungunya IgM		IgM - rapid test	CDC: Low sensitivity (Johnson et al., 2016; Prat et al., 2014; Kosasih et al., 2012; Blacksell et al., 2011)

in antibody titers in paired specimens (collected during the acute and convalescent phases of the disease) are generally used as criteria to make a reliable diagnosis of infection.

3.2.3. Commercial serological tests (IgM)

A search for commercially available tests identified different kits dedicated to the detection of anti-CHIKV IgM, most of which have been evaluated in clinical or EQA studies (Johnson et al., 2016; Prat et al., 2014; Kosasih et al., 2012; Yap et al., 2010; Litzba et al., 2008; Prince et al., 2016; Blacksell et al., 2011) (Supplementary data-Table 2 and Table 3). Immunochromatographic rapid tests (CTK and SD Diagnostics) are an attractive diagnostic option, but their performances are characterized by low sensitivity (Johnson et al., 2016; Prat et al., 2014; Kosasih et al., 2012; Yap et al., 2010; Blacksell et al., 2011). IFT by EUROIMMUN proved to be highly sensitive and specific (96,9% and 98,3% respectively) (Litzba et al., 2008), although a variation in sensitivity was observed in two independent outbreaks caused by different strains of CHIKV, possibly due to amino acid differences in the viral E1 and E2 envelope proteins (Yap et al., 2010). Immuno-enzymatic tests represent the most common commercially available assays, although performances in terms of sensitivity and specificity are quite diversified. According to Johnson et al. (2016), who evaluated most of them, IgM ELISA by EUROIMMUN and InBios have the highest sensibility and sensitivity (although in a follow-up study EUROIMMUN ELISA was affected by cross-reaction, with detection of anti-ONNV IgM (Prat et al., 2014)). The Abcam ELISA also gave reliable results, although a considerable batch-to-batch variability was observed; ELISA assays by CTK, Genway, SD and IBL lacked sensitivity and are not recommended in their current format.

3.2.4. Commercially available serological tests (IgG)

Some of the commercially available tests for IgG detection have been evaluated (Supplementary data-Table 3 and Table 4). IFT (immunofluorescent test) by EUROIMMUN showed high sensitivity and specificity (95,4% and 100%, respectively) (Litzba et al., 2008). ELISA by EUROIMMUN and IBL proved to be specific (95% and 96%), but had lower sensitivity (88% and 52%); both detected anti-ONNV IgG, and the EUROIMMUN ELISA also anti-MAYV IgG (Prat et al., 2014).

3.2.5. International standard (IS)

No IS for serological test is currently distributed.

3.2.6. External quality assessment (EQA)

Three EQAs (Table 5) were organized by ENIVD to investigate laboratories' capacity for serological detection of CHIKV infections. Donoso-Mantke organized in 2007 an EQA analysis that included 24 laboratories from 15 European countries (Donoso Mantke and Niedrig, 2007). Of 18 participants that performed serologic assays, 14 tested for both IgM and IgG and 4 tested only for IgG. Serology testing revealed greater differences amongst laboratories than molecular testing. Good performances were proved for 8 out of 14 laboratories testing for both IgM and IgG, as well as for 3 out of 4 laboratories testing only for IgG; all the other laboratories lacked in sensitivity. No false-positive reactions from cross-reactivity with antibodies against viruses other than CHIKV were observed.

Table 4

Performances of commercial tests for detection of anti-CHIKV IgG, according to published evaluations.

Kit	Company	Method	Performance (ref)
Anti CHIKV IFT	EUROIMMUN	IFT	Good sensitivity and specificity; Variation in sensitivity for Ab against different CHIKV strains (Yap et al., 2010; Litzba et al., 2008)
Anti-CHIKV ELISA (IgG)		IgG ELISA	Quite good sensitivity and specificity; Cross-reactivity with anti-ONNV and anti-MAYV Abs; False neg and false pos results (Prat et al., 2014)
CHIKV Detect TM IgG ELISA Kit	InBios	IgG ELISA	Quite good sensitivity and specificity; Cross-reactivity with anti-ONNV Abs; False neg and false pos results (Prat et al., 2014)

In 2007, 30 expert laboratories from 23 countries in Europe, the Middle East, Asia, Africa, North America and the Caribbean were involved in the second EQA on diagnostic serological proficiency (Niedrig et al., 2009). Only 6/30 obtained the highest score: for the others, a lack of sensitivity, especially for IgM, was observed. Most laboratories used in-house tests; IFT IgM/IgG by EUROIMMUN was the most common commercial assay. No significant variation in performance was observed when comparing the assay type (immunoenzymatic *versus* immunofluorescence assays) or origin (in-house *versus* commercial assays); a strong variability in diagnostic accuracy was reported among laboratories using the same commercial assay, probably due to improper handling of samples and/or assays.

The most recent EQA for serological detection of CHIKV was organized in 2014 (Jacobsen et al., 2016) involving 56 laboratories from 40 countries in Europe, Asia, Africa, the Middle East, the Americas, the Caribbean and Oceania; 46 and 50 data sets were returned for anti-CHIKV IgG and IgM, respectively. A lack of sensitivity and, to a lesser extent, specificity, were more common for IgM detection than for IgG, with 1/50 and 20/46 laboratories achieving the highest score for IgM and IgG, respectively. The most widely used type of technology was a commercial IFT, followed by in-house ELISA, commercial ELISA and in-house IFT; other techniques such as virus neutralization test (VNT) and haemagglutination inhibition assays (HI) were rarely used. Commercial IFT assays were less capable of detecting low IgM titres, but not with IgG in the same dilution series; in-house ELISAs proved to be more sensitive than commercial ELISAs, but less sensitive than IFT and VNT.

3.2.7. Cryoglobulinaemia (IgM)

Cryoglobulins are single or mixed immunoglobulins that undergo reversible precipitation at low temperatures. Cryoglobulinaemia refers to a condition with cryoglobulins in the serum; it has been described for several infectious diseases, particularly hepatitis C infection. A high prevalence of CHIKV-mixed cryoglobulinaemia (MC) (with type II, II-III or III cryoglobulins) has been described by Oliver et al. (2009) in CHIKV-infected travellers coming back from the Western Indian Ocean. According to this study, CHIKV-MC can lead to misdiagnosis of the disease when ELISAs are performed on samples kept at 4 °C: specific anti-CHIKV IgM could be trapped in the cryoprecipitate, causing unexpected seronegativity for patients with clinical suspicion of CHIKV infection. To circumvent the problem, it is suggested to manage blood samples as required for any cryoglobulin research: sampling and centrifugation at 37 °C, decantation and serum pre-warming before the ELISA assays.

3.2.8. Cross-reactivity

Because of phylogenetic relationships among the three viruses, cross-reactivity, especially between anti-CHIKV and anti-ONNV antibodies, is a major concern when serological tests are performed to make a reliable diagnosis. Cross-reactivity between anti-ONNV and anti-CHIKV antibodies has been primarily investigated as MAYV had a distinct geographic distribution in the past; however, with the appearance of CHIKV in the Americas (Leparc-Goffart et al., 2014) and MAYV in the Caribbean (Lednicky et al., 2016), the antigenic relationship between the two viruses needs to be evaluated more carefully.

CHIKV is closer phylogenetically to ONNV than to MAYV; this

Table 5
EQAs for serological detection of CHIKV.

Year	Reference	Participants	Study coordinator	Objectives	Other viruses tested in the panels	Results of serological assessment
2007	Donoso Mantke and Niedrig (2007)	Total: 24 labs from 15 European countries/ 18 labs for CHIKV serology	European Network for Diagnostics of Imported Viral Diseases (ENIVD)	Assessing quality for CHIKV molecular and serological diagnostics	–	8/14 labs showed good results for both IgM and IgG
2007	Niedrig et al. (2009)	Total: 30 labs from 23 countries (Europe, Middle East, Asia, Africa, North America, Caribbean)	European Network for Diagnostics of Imported Viral Diseases (ENIVD)	Assessing quality for CHIKV serological diagnostics	DENV, WNV	6/30 labs with 100% correct results, 13/30 labs with $\geq 85\%$ correct results, 11/30 labs with $\leq 75\%$ correct results
2014	Jacobsen et al. (2016)	Total: 56 labs from 40 countries (Europe, Asia, Africa, Middle East, Americas, Caribbean, Oceania)	European Network for Diagnostics of Imported Viral Diseases (ENIVD)	Assessing quality for CHIKV molecular and serological diagnostics	RRV, WNV, DENV	IgM: 1/50 dataset classified as “optimal”, 9/50 “acceptable”, 40/50 “need for improvement”. IgG: 20/46 “optimal”, 2/46 “acceptable”, 24/46 “need for improvement”

explains the substantial cross-reactivity observed for both IgM and IgG using ELISA and IFT (Eastwood et al., 2017; Calisher et al., 1986; Prat et al., 2014; Smith et al., 2018). However, there are no documented studies evaluating the exact incidence of serologic cross-reactivity between the two viruses. Even in seroneutralization, it is difficult to differentiate antibodies against CHIKV and ONNV unless consistent differences in reciprocal cross-neutralization occur, which is not typical.

CHIKV and MAYV are more distant phylogenetically, but cross-reactivity has been observed with both ELISA and IFT (Calisher et al., 1986; Prat et al., 2014; Hassing et al., 2010); presumably, it is less extensive than with ONNV, but no accurate estimation of cross-reactivity incidence between CHIKV and MAYV could be identified in the literature.

3.2.9. Cross-neutralization and cross-protection

CHIKV-ONNV-MAYV: Cross-neutralization and cross-protection are important aspects to consider, because of implications for disease spread, as well as from the perspective of vaccine development. Despite this, few studies have been performed, suggesting that anti-CHIKV antibodies can neutralize and protect against MAYV and ONNV infections better than how anti-ONNV and anti-MAYV antibodies can do against CHIKV infection.

One-way cross-neutralization has been demonstrated for CHIKV-ONNV, with anti-CHIKV immune serum inhibiting ONNV plaque formation, while antiserum to ONNV is less effective against CHIKV strains (Chanas et al., 1979). Similar results were observed in studies showing that serum or monoclonal antibodies (mabs) derived from ONNV-infected animals or humans weakly neutralize CHIKV (Porterfield, 1961; Blackburn et al., 1995): Blackburn et al. observed that, using an immunofluorescent test (IFT), 86% of the mabs against CHIKV reacted with ONNV, whereas only 53% of the ONNV mabs reacted with CHIKV strain. A possible explanation could be that, during its evolution, ONNV has retained most of the CHIKV antigenic sites, whereas some of ONNV epitopes have undergone greater conformational change, so that mabs prepared against them neutralize weakly or not at all against CHIKV (Blackburn et al., 1995).

The ability of anti-CHIKV antibodies to neutralize and protect against ONNV infection has been recently investigated *in vivo*: a recombinant CHIKV candidate vaccine was demonstrated to elicit a strong cross-neutralizing antibody response in a mouse model, conferring protection also against ONNV infection (Partidos et al., 2012).

As concerns cross-neutralization between CHIKV and MAYV, a plaque-neutralization test has showed that MAYV antiserum slightly neutralizes CHIKV, while no inhibition effect on MAYV has been caused by CHIKV antiserum (Porterfield, 1961).

Considering overall findings about anti-CHIKV antibodies effect on MAYV and ONNV, they are diversified and need further assessment. Porterfield et al. (Porterfield, 1961) showed that CHIKV antiserum neutralizes efficiently ONNV but not MAYV. These results contrast with what observed by Fox et al. (2015) using two murine mabs against epitopes on the B domain of the CHIKV E2 protein: *in vitro*, they cross-neutralized MAYV more than ONNV; in a mouse model, they proved to be able to protect against MAYV infection, and to reduce disease caused by ONNV.

4. O'nyong-nyong and Mayaro virus

4.1. Molecular diagnosis

4.1.1. Kinetics of viral load and dedicated biological sampling

ONNV: Only few published studies have examined ONNV kinetics; similarly to CHIKV, ONNV viremia seems to last approximately one week or less (Kiwanuka et al., 1999).

MAYV: Case reports suggest a very narrow window during which molecular assays can detect circulating virus. Halsey et al. (2013) observed negative results in two samples collected beyond day 3 of

symptoms and tested by PCR; similarly, 4 days after the onset of illness Coimbra et al. could not detect viremia from two MAYV-case blood samples (Coimbra et al., 2007). However, an extended viremia (10 days) has been observed in an HIV-infected patient, possibly due to his immunocompromised status (Estofotele et al., 2016).

For both ONNV and MAYV, molecular tests were performed only on sera collected at the acute stage of infection.

4.1.2. Choice of test

ONNV – MAYV: Suggestions about the choice of test for molecular diagnosis of infections are the same provided to detect CHIKV infection and have been discussed in the paragraph 3.1.2.

4.1.3. In house real-time RT-PCR

ONNV – MAYV: Few published studies have described PCR platforms for the detection of ONNV (Smith et al., 2009; Waggoner et al., 2017a; Liu et al., 2016; Saxton-Shaw et al., 2013); they are listed in Table 6. As concerns MAYV, an alignment of 20 sequences available in GenBank, representing the main MAYV lineages, was made to assess primer sets developed in-house by various laboratories. The results of this *in silico* analysis are presented in Table 7; as explained for CHIKV, a single mutation in the five 3' terminal positions of a primer can significantly compromise its detection capability, so mismatches concerning nucleotides in these positions are highlighted in the table. Most PCR systems use primers and probe showing one or more mismatches with MAYV strains.

4.1.4. Commercial molecular tests

ONNV – MAYV: They are not included as target in any commercial tests.

4.1.5. International standard (IS)

ONNV – MAYV: No international standard for molecular detection of the two viruses has been developed.

4.1.6. External quality assessment (EQA)

ONNV – MAYV: No EQAs have ever been organized to evaluate laboratory' capacity for ONNV and MAYV detection.

4.2. Serological diagnosis

4.2.1. Kinetics of immune response

ONNV: Little is known about the kinetics of the antibody response to ONNV. IgM antibodies typically peak two weeks after the onset of illness and persist for about two months (Bessaud et al., 2006; Kiwanuka et al., 1999), although a few cases have been described with detectable IgM for 6 months or more (Kiwanuka et al., 1999). Information about anti-ONNV IgG comes only from two case reports: a traveller returning from Kenya experienced IgG seroconversion 26 days after disease onset (Tappe et al., 2014); a report from Chad described a peak in a patient's IgG titre 68 days after the acute stage of illness (Bessaud et al., 2006).

MAYV: Anti-MAYV antibody kinetics are also poorly documented. IgM typically appears three days after the onset of illness (Mota et al., 2015) and lasts for three months or more, but not beyond six months (Halsey et al., 2013; Mota et al., 2015). In contrast, IgG may persist for years (Mota et al., 2015). In a case report, the absence of IgG seroconversion was reported in a patient 3 months after onset of illness, probably because the time between disease onset and the last blood sampling in this patient was too short to allow Ig class switching (Receveur et al., 2010).

4.2.2. Choice of test

ONNV-MAYV: Suggestions about the choice of test for serological diagnosis of infections are the same provided to detect CHIKV infection and have been discussed in the paragraph 3.2.2.

4.2.3. Commercial serological tests

ONNV: There is no commercial kit for IgM or IgG detection.

MAYV: EUROIMMUN developed an anti-MAYV ELISA (IgM-IgG) but it has not been evaluated.

4.2.4. International standard (IS)

ONNV-MAYV: No IS for serological tests is currently available.

4.2.5. External quality assessment (EQA)

ONNV-MAYV: No EQA has been organized to evaluate laboratory capacity for the serological diagnosis of ONNV or MAYV infection.

4.2.6. Cryoglobulinaemia

ONNV-MAYV: The presence of cryoglobulins in sera from patients positive for CHIKV infection should have encouraged to assess it for ONNV and MAYV also. Actually, no study about cryoglobulinaemia in sera from patients positive for ONNV or MAYV has been performed.

4.2.7. Cross-reactivity, cross-neutralization and cross-protection

These items have been discussed in the paragraph 3.2.8 and 3.2.9.

5. Discussion

The development, validation and evaluation of diagnostic tools are crucial steps to developing accurate diagnostic methods of alphavirus infections. A considerable amount of financial resources and efforts are necessary to implement virus and virus-specific antibody detection. This justifies why diagnostic tools ideally should be developed before outbreaks, in order to ensure a rapid response, and it's valid especially for pathogens showing clear assumptions of large-scale dissemination.

The recent occurrence of extensive CHIKV epidemics has necessitated improved documentation of infection and has impelled laboratories and companies to develop specific molecular and serological assays. However, it has also highlighted the need for improved capacity for diagnostic surveillance, especially when the co-circulation of closely related viruses increases the chance of misdiagnosis.

ONNV and MAYV have been discussed in parallel with CHIKV because their similar clinical presentations, serological cross-reactivity and geographic areas of co-circulation are significant barriers to specific diagnosis.

6. Knowledge gaps

6.1. Chikungunya virus

- Kinetics of viral load and dedicated biological sampling: different kinds of clinical samples have been used for the molecular detection of the virus. However, for some of them (i.e. urine and breast milk) the time window for detection of the viral genome remains unclear, with different authors reporting contradictory results. Comparative

Table 6

Published reports of RT-PCR systems for the detection of ONNV, in decreasing order of the year of publication.

Year	Reference	Method	Target
2017	Waggoner et al. (2017a)	Monoplex real time RT-PCR	5'UTR-nsP1
2016	Liu et al. (2016)	Real-time PCR-based TaqMan array card (TAC) (15 viruses)	E1
2013	Saxton-Shaw et al. (2013)	Monoplex real time RT-PCR	E1
2009	Smith et al. (2009)	Monoplex real time RT-PCR	E1 nsP1 nsP2
		Multiplex (ONNV-CHIKV) real time RT-PCR	5'UTR-nsP1

Table 7

We selected ten sequences for MAYV lineage D (GenBank accession numbers: KP842806, KP842795, KM400591, KJ013266, DQ001069, KP842807, KP842813, KP842800, KP842802, KP842809), one for MAYV lineage N (GenBank accession number: KP842812), nine for MAYV lineage L (GenBank accession numbers: KP842819, KP842820, KP842818, KT818520, KX496990, AF237947, NC.003417, KY618133, KY985361). Mismatch positions between this reference sequence panel and the primers and probes are indicated in the table; mismatches concerning the five 3' terminal nucleotides of a primer are indicated in red and bold characters. In probes, 1 or 2 mutations do generally not compromise the hybridization potential but this also depends on the length of the probe and its G + C content. D: dispersed; L: limited; N: new.

Target(s)	Reference	Technique	Amplicon size (bp)	FW primer 5'-3'	RV primer 5'-3'	Probe 5'-3'	MAYV D (nb seq = 10)	MAYV N (nb seq = 1)	MAYV L (nb seq = 9)
5'UTR-nsP1	Waggoner et al. (2017b)	Hydrolysis probe	109	AAGCTCTTCCTCTGCAITGG	TGCTGGAAACYGCTCTYTGTA	CCCGAGAGCCGGTTTTTAAATCAG	Pos 14: T > A (1)		
nsP1	Naveca et al. (2017)	Hydrolysis probe	59	CACGGACMTTITGGCTTCA	AGACTGCCACCTCTGCTKAGAG	Pos 2: A > G (1) Pos 8: T > 1 (1) Pos 17: T > C (2)	Pos 11: C > T (1)	Pos 17: T > C (1)	T > C (3)
				ACAGATCAGATGCAGG	Pos 3: T > C (1)				
E1	Llagonne-Barets et al. (2016)	SYBR green	165	TTCGAAAYCAAGTGGGATTC	NA				
nsP1	Friedrich-Jänicke et al. (2014)	Hydrolysis probe	95	GCCTGGAAGTACAAGAA	GGTGGCAGTGTATCAGGATGTCTATG	A > G (2) G > A (5) C > T (2)	G > A (1)	A > G (1) G > A (1)	Pos 1: G > C (2) Pos 8: A > T (3), A > C (5) Pos 13: T > C (4) Pos 12: A > C (9) Pos 21: T > C (8) C > T (6) T > C (9) A > T (5)
				CAAATGTCACACAGGCGAAG	Pos 8: A > G (1)				
E2	Long et al. (2011)	Hydrolysis probe	127	GTGGTGGCAGAGTGAATCTTTC	GACCTGTGGGATAGCCTACCCACCAT	Pos 12: A > C (4) Pos 15: C > T (1) C > T (1)	Pos 9: G > A (1)	Pos 9: G > A (1) C > T (1)	

studies to assess body fluids are missing: they should be performed to better document kinetics of viral loads and to identify the most appropriate samples for diagnostic use at the different steps of the disease.

- In-house real-time PCR protocols: several in-house real-time PCR protocols have been published, with different performances in the detection of the CHIKV lineages predicted from *in silico* analyses. An *in vitro* analytical evaluation would allow to better assess the adequacy of PCR systems to detect different lineages.
- Commercial real-time PCR assays: a large number of commercial tests are available for molecular diagnosis. The performances of only one kit have been reported under the form of a scientific article. Kits datasheets commonly provide insufficient information regarding the lineage(s) detected and how the performances of the assay were evaluated. Lack of information about primers and probes used does not allow assessment of adequacy for covering existing genetic variability. In the context of External Quality Assessments, some commercial tests are not clearly identified.
- Molecular International Standards: development of IS is essential for harmonisation of results among different laboratories; however, the only IS available for molecular diagnosis has been prepared with one ECSA strain, despite the significant genetic diversity between CHIKV lineages.
- Commercially available serological assays: a large number is available and several evaluations have been published; however, end-users would benefit from the guidance of a global and independent test evaluation.

6.2. *O'nyong-nyong* and *Mayaro* virus

- Kinetics of viral load and dedicated biological sampling: few cases are described in the literature and do not provide a clear picture of the kinetics of viremia. This could be problematic especially for MAYV, which may have a shorter window of viremia compared to CHIKV and ONNV. To the best of our knowledge, no samples other than sera have been tested, so that kinetics of viral loads in different body fluids have never been described.
- Kinetics of the immune response: it is only documented from a few case reports; accordingly, it is not clear when antibodies would appear and how long they would persist in sera.
- In-house real-time PCR protocols: a few in-house real-time PCR protocols have been published for both pathogens. *In silico* analysis shows that most primers and probe sets have one or more mismatches with some MAYV evolutionary lineages. An *in vitro* analytical evaluation would allow to better assess the performances of the PCR systems and their adequacy to detect the different MAYV lineages.
- Commercial molecular and serological tests: none is available for ONNV; one serological assay (IgM) is available for MAYV, but it has not been evaluated.
- Molecular International Standards: none is available.
- External Quality Assessments (EQAs): they have never been organized to evaluate laboratories' capacity to diagnose ONNV and MAYV infection.

6.3. All three viruses

- Cryoglobulinaemia: false negative serological results due to cryoglobulinaemia have been reported for patients infected with CHIKV (only one study available). The presence and role of cryoglobulinaemia have not been investigated for ONNV and MAYV.
- Cross reactivity and virus co-circulation: substantial cross-reactivity between anti-CHIKV and anti-ONNV antibodies, as well as between anti-CHIKV and anti-MAYV has been reported, but poorly characterized. Differential diagnosis (CHIKV-ONNV and CHIKV-MAYV) in co-circulation areas is a challenging task. Serological assays

allowing differential identification of the infection by the different viruses are required.

- Cross-neutralization and cross-protection: few studies have been performed, which limits the ability to assess cross-neutralization and/or cross-protection.

7. Expert recommendations

7.1. *Chikungunya* virus

- In-house real-time PCR protocols: the numerous laboratory-developed protocols would benefit from a shared platform with *in silico* analysis of primers and probe sets; it could help to identify the most appropriate PCR for detection of circulating strains. A similar service has been previously proposed by the European Virus Archive website upon emergence of the CHIKV Asian genotype in the Caribbean. A host website governed by a non-commercial academic or public health organization, should be identified to implement this database and to contain results of laboratory comparative tests. After having defined mechanism of analysis and collection of results, this could be proposed as a model for other emerging viral diseases.
- Commercially available real-time PCR tests (i): analytical evaluation and comparative studies of commercial molecular assays should be performed to assess both sensitivity and specificity. Results of External Quality Assessments should be made available with a clear identification of the commercial assays used.
- Commercially available real-time PCR tests (ii): to assess the adequacy of commercial tests for genomic diversity of CHIKV, companies should make publicly available detailed information about primers and probes sets and/or provide updated performance analysis of their kits for detection of existing and newly identified variants.
- Serological tests: to better assess laboratory capacity for serological diagnosis, reference panels should be constituted and tested using operational tests. This could help to evaluate performances of both commercial and in-house tests.
- Commercially available serological tests: including commercial tests in EQAs and in comparative studies can help laboratories in their choice for the most reliable diagnostic assays.

7.2. *O'nyong-nyong* and *Mayaro* virus

- In general: both ONNV and MAYV require specific efforts for implementing clinical and epidemiological studies.
- Viremia: viremia kinetics, as well as viral loads in different body fluids, should be better documented.
- Immune response: the kinetics of the antiviral IgM and IgG response should be evaluated during the course of natural infection.
- Commercial molecular and serological tests: they should be developed and evaluated through comparative studies.
- International Standards (IS): they should be made available, taking into account the MAYV genetic heterogeneity.
- External Quality Assessments (EQAs): they should be organized to assess laboratory capacity of detecting ONNV and MAYV with molecular and serological tools.

7.3. All three viruses

- Kinetics of viral load and dedicated biological sampling: a detailed analysis of the presence of the viruses in different body fluids over the course of infection should be performed and lead to a rational standardization of the process of clinical sampling according to the clinical presentation and stage of the disease.
- In-house real-time PCR protocols: the lack of information about assay performances could be filled by organizing comparative studies, with several laboratories testing reference panels by molecular

methods; the evaluation should include other viruses at risk of cross-reactivity and different CHIKV and MAYV genotypes, in order to determine limit of detection (LOD) for each genotype. If necessary, new protocols for molecular diagnosis should be developed.

- Genomic sequence database: experts recommend that a genomic reference database should be made available similar to those existing for other viruses (e.g. see the sites of the Virus Variation Resource ([Virus Variation](#)) or the Virus Pathogen Resource ([Virus Pathogen Database and Analysis Resource](#)). This database would allow to store available sequence data, together with gene and protein annotations and information about isolation hosts and sources.
- International Standards (IS): it is highly recommended to develop IS for CHIKV Asian and West-African lineages, as well as for MAYV and ONNV.
- Cryoglobulinaemia: further studies should be implemented to confirm and better document the impact of cryoglobulinaemia in unexpected CHIKV seronegativity, as well as to extend investigations to ONNV and MAYV infections.
- Co-circulation: viral co-circulation requires the development of molecular and serological multiplex tools to differentiate CHIKV from ONNV in Africa and from MAYV in Southern and Central America.
- Cross-reactivity: multidirectional studies should be implemented to better define the exact extent of CHIKV cross-reactivity with ONNV and MAYV; this would be most probably required obtaining characterized samples for both naturally exposed humans and experimentally infected non-human primates.
- Cross-neutralization and cross-protection: cross-neutralization and cross-protection studies between the three viruses should be implemented, including studies to identify potential ONNV- and MAYV- specific monoclonal antibodies; as for cross-reactivity, samples from both naturally exposed humans and experimentally infected non-human primates should prove useful for this purpose.

In summary, analysis of the currently available literature and consultation of experts indicate that inadequate diagnostics are currently available for ONNV and MAYV and, to a lesser degree, for CHIKV. Gaps in diagnostic tools and protocols allowing the identification of the etiological agent during the course of CHIKV, ONNV and MAYV infections have been identified for both diagnosis during the acute phase and for the long-term follow-up of patients, as well as for sero-epidemiological studies.

Diagnosis of the disease at the acute phase relies mostly on molecular detection of the virus genome. Molecular assays have been described for CHIKV, ONNV and MAYV but require further evaluation, standardized protocols and the availability of international standards representing the genetic diversity of the viruses. Alternatively, diagnosis can be provided by the detection of specific IgM, but the exact extent of cross-reactivity between the three viruses, the sensitivity of the assays, and the possible interfering role of cryoglobulinaemia require further investigation. Implementation of reference panels and EQAs for both molecular and serological assays is necessary.

Regarding sero-epidemiological studies, there is no reported high throughput assay that enables the different viruses to be distinguished in areas of potential co-circulation. Even neutralization assays can lead to ambiguous interpretation. New specific tools and/or improved standardized protocols are needed to enable large-scale epidemiological studies of public health relevance to be performed.

Regarding the long-term follow-up of patients infected by CHIKV (and potentially ONNV and MAYV) there is currently no biological marker of medical significance associated with disease progression and prognosis. This obviously requires a complete reassessment and specific investigation.

Finally, when pathogens emerge or re-emerge, there is little time for development of diagnostics, which should be designed and validated in

advance to ensure a rapid response. Considering the high risk of future CHIKV, MAYV and ONNV outbreaks, it is highly recommended that a major investigation should be initiated to fill existing diagnostic gaps.

Acknowledgement

Publication fees were covered by the Global Research Collaboration for Infectious Disease Preparedness (GloPID-R).

Appendix A. Supplementary data

Supplementary data to this article can be found online at <https://doi.org/10.1016/j.antiviral.2019.03.009>.

References

- Althouse, B.M., Guerbois, M., Cummings, D.A.T., Diop, O.M., Faye, O., Faye, A., et al., 2018 13. Role of monkeys in the sylvatic cycle of chikungunya virus in Senegal. *Nat. Commun.* 9 (1), 1046.
- Anderson, C.R., Downs, W.G., Wattley, G.H., Ahin, N.W., Reese, A.A., 1957 Nov 1. Mayaro virus: a new human disease agent. *Am. J. Trop. Med. Hyg.* 6 (6), 1012–1016.
- Añez, G., Heisey, D.A., Rios, M., 2014 Jul 31. Complete coding region sequence of a chikungunya virus strain used for formulation of CBER/FDA RNA reference reagents for nucleic acid testing. *Genome Announc.* 2 (4) [Internet] [cited 2018 Feb 20]. Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC4118057/>.
- Añez, G., Jiang, Z., Heisey, D. a. R., Kerby, S., Rios, M., 2015 Nov 1a. The Chikungunya virus Collaborative Study Group. Collaborative study for the characterization of a chikungunya virus RNA reference reagent for use in nucleic acid testing. *Vox Sang.* 109 (4), 312–318.
- Añez, G., Jiang, Z., Heisey, D. a. R., Kerby, S., Rios, M., 2015 Novb. Chikungunya virus Collaborative Study Group. Collaborative study for the characterization of a chikungunya virus RNA reference reagent for use in nucleic acid testing. *Vox Sang.* 109 (4), 312–318.
- Appassakij, H., Khuntikij, P., Kemapunmanus, M., Wutthanarungsan, R., Silpapojakul, K., 2013 Oct. Viremic profiles in asymptomatic and symptomatic chikungunya fever: a blood transfusion threat? *Transfusion (Paris)* 53 (10 Pt 2), 2567–2574.
- Bandeira, A.C., Campos, G.S., Rocha, V.F.D., Souza, B.S. de F., Soares, M.B.P., Oliveira, A.A., et al., 2016. Prolonged shedding of Chikungunya virus in semen and urine: a new perspective for diagnosis and implications for transmission. *IDCases* 6, 100–103.
- Bessaud, M., Peyrefitte, C.N., Pastorino, B.A.M., Gravier, P., Tock, F., Boete, F., et al., 2006 Aug. O'nyong-nyong Virus. *Chad. Emerg. Infect. Dis.* 12 (8), 1248–1250.
- Blackburn, N.K., Besselaar, T.G., Gibson, G., 1995 Feb. Antigenic relationship between chikungunya virus strains and o'nyong nyong virus using monoclonal antibodies. *Res. Virol.* 146 (1), 69–73.
- Blacksell, S.D., Tanganuchitcharnchai, A., Jarman, R.G., Gibbons, R.V., Paris, D.H., Bailey, M.S., et al., 2011 Oct. Poor diagnostic accuracy of commercial antibody-based assays for the diagnosis of acute chikungunya infection. *Clin. Vaccine Immunol.* CVI 18 (10), 1773–1775.
- Borgherini, G., Poubreau, P., Staikowsky, F., Lory, M., Moullec, N.L., Becquart, J.P., et al., 2007. Outbreak of chikungunya on reunion island: early clinical and laboratory features in 157 adult patients. *Clin. Infect. Dis.* 44 (11), 1401–1407.
- Bozza, F., Moreira-Soto, A., Rockstroh, A., Fischer, C., Calheiros, A., Drosten, C., et al., 2019 Feb. Differential shedding and antibody kinetics of Zika virus and chikungunya virus, Brazil. *Emerg. Infect. Dis* [Internet] Available from: <https://doi.org/10.3201/eid2502.180166>.
- Calba, C., Guerbois-Galla, M., Franke, F., Jeannin, C., Auzet-Cailaud, M., Grard, G., et al., 2017. Preliminary report of an autochthonous chikungunya outbreak in France, July to September 2017. *Euro Surveill.* 22 (39) 17–00647.
- Calisher, C.H., el-Kafrawi, A.O., Al-Deen Mahmud, M.I., Travassos da Rosa, A.P., Bartz, C.R., Brummer-Korvenkontio, M., et al., 1986 Jan. Complex-specific immunoglobulin M antibody patterns in humans infected with alphaviruses. *J. Clin. Microbiol.* 23 (1), 155–159.
- Campos, G.S., Bandeira, A.C.A., Rocha, V.F.D., Dias, J.P., Carvalho, R.H., Sardi, S.I., 2017. First detection of chikungunya virus in breast milk. *Pediatr. Infect. Dis. J.* 1.
- Carrera, J.-P., Bagamian, K.H., Rosa, A.P.T. da, Wang, E., Beltran, D., Gundaker, N.D., et al., 2018 Jun 6. Human and equine infection with alphaviruses and flaviviruses in Panama during 2010: a cross-sectional study of household contacts during an encephalitis outbreak. *Am. J. Trop. Med. Hyg.* 98 (6), 1798–1804.
- Casolari, S., Briganti, E., Zanotti, M., Zauli, T., Nicoletti, L., Magurano, F., et al., 2008. A fatal case of encephalitis associated with Chikungunya virus infection. *Scand. J. Infect. Dis.* 40 (11–12), 995–996.
- Cecilia, D., Kakade, M., Alagarasu, K., Patil, J., Salunke, A., Parashar, D., et al., 2015. Development of a multiplex real-time RT-PCR assay for simultaneous detection of dengue and chikungunya viruses. *Arch. Virol.* 160 (1), 323–327.
- Chanas, A.C., Hubalek, Z., Johnson, B.K., Simpson, D.I., 1979. A comparative study of O'nyong nyong virus with Chikungunya virus and plaque variants. *Arch. Virol.* 59 (3), 231–238.
- Chandak, N.H., Kashyap, R.S., Kabra, D., Karandikar, P., Saha, S.S., Morey, S.H., et al., 2009 Jan 3. Neurological complications of Chikungunya virus infection. *Neurol. India* 57 (2), 177.
- Chikungunya. World Health Organization. [Internet] [cited 2018 Nov 5]. Available from: <http://www.who.int/news-room/fact-sheets/detail/chikungunya>.
- Chiu, C.Y., Bres, V., Yu, G., Krysztof, D., Naccache, S.N., Lee, D., et al., 2015 Aug. Genomic assays for identification of chikungunya virus in blood donors, Puerto Rico,

2014. *Emerg. Infect. Dis.* 21 (8), 1409–1413.
- Chow, A., Her, Z., Ong, E.K.S., Chen, J., Dimatatac, F., Kwek, D.J.C., et al., 2011 Jan 15. Persistent arthralgia induced by Chikungunya virus infection is associated with interleukin-6 and granulocyte macrophage colony-stimulating factor. *J. Infect. Dis.* 203 (2), 149–157.
- Chusri, S., Siripaitoon, P., Silpapojakul, K., Hortiwakul, T., Charernmak, B., Chinnawirotpisan, P., et al., 2014 Mar 5. Kinetics of chikungunya infections during an outbreak in Southern Thailand, 2008–2009. *Am. J. Trop. Med. Hyg.* 90 (3), 410–417.
- Cleton, N., Koopmans, M., Reimerink, J., Godeke, G.J., Reusken, C., 2012. Come fly with me: review of clinically important arboviruses for global travelers. *J. Clin. Virol.* 55 (3), 191–203.
- Coffey, L.L., Failloux, A.-B., Weaver, S.C., 2014 Nov 24. Chikungunya virus–vector interactions. *Viruses* 6 (11), 4628–4663.
- Coimbra, T.L.M., Santos, C.L.S., Suzuki, A., Petrella, S.M.C., Bisordi, I., Nagamori, A.H., et al., 2007 Aug. Mayaro virus: imported cases of human infection in São Paulo State. *Brazil. Rev. Inst. Med. Trop. São Paulo* 49 (4), 221–224.
- Dengue Fever PCR Test | Multiplex PCR Test | Fast-Track Diagnostics [Internet]. [cited 2018 Dec 5]. Available from: <http://www.fast-trackdiagnostics.com/human-line/products/ftd-denguechik/>.
- Diagnostic Testing, 2017. Chikungunya Virus|CDC. [Internet] [cited 2018 Nov 14]. Available from: <https://www.cdc.gov/chikungunya/hc/diagnostic.html>.
- Diallo, M., Thonnon, J., Traore-Lamizana, M., Fontenille, D., 1999 Feb. Vectors of Chikungunya virus in Senegal: current data and transmission cycles. *Am. J. Trop. Med. Hyg.* 60 (2), 281–286.
- Donoso Mantke, O., Niedrig, M., ENIVD members, 2007 Sep 13. Laboratory capacity for detection of chikungunya virus infections in Europe. *Euro Surveill. Bull. Eur. Sur Mal Transm. Eur. Commun. Dis. Bull.* 12 (9) E070913.2.
- Eastwood, G., Sang, R.C., Guerbois, M., Taracha, E.L.N., Weaver, S.C., 2017 Nov 8. Enzootic circulation of chikungunya virus in east Africa: serological evidence in non-human Kenyan primates. *Am. J. Trop. Med. Hyg.* 97 (5), 1399–1404.
- Edwards, C.J., Welch, S.R., Chamberlain, J., Hewson, R., Tolley, H., Cane, P.A., et al., 2007 Aug 1. Molecular diagnosis and analysis of Chikungunya virus. *J. Clin. Virol.* 39 (4), 271–275.
- Estofolete, C.F., Mota, M.T.O., Vedovello, D., Góngora, D.V.N. de, Maia, I.L., Nogueira, M.L., et al., 2016 Oct. Mayaro fever in an HIV-infected patient suspected of having Chikungunya fever. *Rev. Soc. Bras. Med. Trop.* 49 (5), 648–652.
- Fox, J.M., Long, F., Edeling, M.A., Lin, H., van Duijl-Richter, M.K.S., Fong, R.H., et al., 2015 Nov 19. Broadly neutralizing alphavirus antibodies bind an epitope on E2 and inhibit entry and egress. *Cell* 163 (5), 1095–1107.
- Friedrich-Jänicke, B., Emmerich, P., Tappe, D., Günther, S., Cadar, D., Schmidt-Chanasit, J., 2014 Jul. Genome analysis of Mayaro virus imported to Germany from French Guiana. *Emerg. Infect. Dis.* 20 (7), 1255–1257.
- Gallian, P., de Lamballerie, X., Salez, N., Piorkowski, G., Richard, P., Patrel, L., et al., 2014 Jun 5. Prospective detection of chikungunya virus in blood donors, Caribbean 2014. *Blood* 123 (23), 3679–3681.
- Gérardin, P., Couderc, T., Bintner, M., Tournebise, P., Renouil, M., Lémant, J., et al., 2016 Jan 5. Chikungunya virus-associated encephalitis: a cohort study on La Réunion Island, 2005–2009. *Neurology* 86 (1), 94–102.
- Grivard, P., Roux, K.L., Laurent, P., Fianu, A., Perrau, J., Gigan, J., et al., 2007. Molecular and Serological Diagnosis of Chikungunya Virus Infection. /data/revues/03698114/00550010/07001174/. <http://www.em-consulte.com/en/article/68222> Dec 18 [cited 2018 Feb 21]; Available from: <http://www.em-consulte.com/en/article/68222>
- Haddow, A.J., Davies, C.W., Walker, A.J., 1960 Nov 1. O'nyong-nyong fever: an epidemic virus disease in East Africa 1. introduction. *Trans. R. Soc. Trop. Med. Hyg.* 54 (6), 517–522.
- Halsey, E.S., Siles, C., Guevara, C., Vilcarromero, S., Jhonston, E.J., Ramal, C., et al., 2013 Nov. Mayaro virus infection, amazon basin region, Peru, 2010–2013. *Emerg. Infect. Dis.* 19 (11), 1839–1842.
- Hassing, R.-J., Lepar-Goffart, I., Blank, S.N., Thevarayan, S., Tolou, H., van Doornum, G., et al., 2010 Oct. Imported Mayaro virus infection in The Netherlands. *J. Infect.* 61 (4), 343–345.
- Heath, C.J., Lowther, J., Noël, T.P., Mark-George, I., Boothroyd, D.B., Mitchell, G., et al., 2018 Jan 1. The identification of risk factors for chronic chikungunya arthralgia in Grenada, west indies: a cross-sectional cohort study. *Open Forum Infect. Dis.* 5 (1) [Internet] [cited 2018 Dec 24]. Available from: <https://academic.oup.com/ofid/article/5/1/ofx234/4788106>.
- Hoch, A.L., Peterson, N.E., LeDuc, J.W., Pinheiro, F.P., 1981 May. An outbreak of Mayaro virus disease in Belterra, Brazil. III. Entomological and ecological studies. *Am. J. Trop. Med. Hyg.* 30 (3), 689–698.
- Hotez, P.J., Murray, K.O., 2017 Aug 31. Dengue, West Nile Virus, Chikungunya, Zika—And Now Mayaro? *PLoS Negl. Trop. Dis.* 11 (8) [Internet] [cited 2018 Nov 5]. Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC5578481/>.
- Jacobsen, S., Patel, P., Schmidt-Chanasit, J., Lepar-Goffart, I., Teichmann, A., Zeller, H., et al., 2016. External quality assessment studies for laboratory performance of molecular and serological diagnosis of Chikungunya virus infection. *J. Clin. Virol.* 76, 55–65.
- Johnson, B.W., Goodman, C.H., Holloway, K., de Salazar, P.M., Valadere, A.M., Drebot, M.A., 2016 Jul 6. Evaluation of commercially available chikungunya virus immunoglobulin M detection assays. *Am. J. Trop. Med. Hyg.* 95 (1), 182–192.
- Jones, P., Okeoma, C., 2015 Aug 8. Detection of chikungunya virus (CHIKV) in urine of infected mice: a potential non-invasive diagnostic tool for CHIKV. *J. Infect. Dis. Ther.* 3 (4), 1–6.
- Kam, Y.-W., Lum, F.-M., Teo, T.-H., Lee, W.W.L., Simarmata, D., Harjanto, S., et al., 2012 Apr. Early neutralizing IgG response to Chikungunya virus in infected patients targets a dominant linear epitope on the E2 glycoprotein. *EMBO Mol. Med.* 4 (4), 330–343.
- Kam, Y.-W., Simarmata, D., Chow, A., Her, Z., Teng, T.-S., Ong, E.K.S., et al., 2012 Apr 1b. Early appearance of neutralizing immunoglobulin G3 antibodies is associated with chikungunya virus clearance and long-term clinical protection. *J. Infect. Dis.* 205 (7), 1147–1154.
- Karabatsos, N., 1975 May. Antigenic relationships of group A arboviruses by plaque reduction neutralization testing. *Am. J. Trop. Med. Hyg.* 24 (3), 527–532.
- Kiwanuka, N., Sanders, E.J., Rwaguma, E.B., Kawamata, J., Ssengooba, F.P., Najjemba, R., et al., 1999 Nov 1. O'Nyong-Nyong fever in South-Central Uganda, 1996–1997: clinical features and validation of a clinical case definition for surveillance purposes. *Clin. Infect. Dis.* 29 (5), 1243–1250.
- Kosasih, H., Widjaja, S., Surya, E., Hadiwijaya, S.H., Butarbutar, D.P.R., Jaya, U.A., et al., 2012 Jan. Evaluation of two IgM rapid immunochromatographic tests during circulation of Asian lineage Chikungunya virus. *Southeast Asian J. Trop. Med. Publ. Health* 43 (1), 55–61.
- LaBeaud, D., Banda, T., Brichard, J., Muchiri, E.M., Mungai, P.L., Mutuku, F.M., et al., 2015. High rates of O'nyong nyong and chikungunya virus transmission in coastal Kenya. *PLoS Negl. Trop. Dis.* 9 (2), 1–14.
- Lanciotti, R.S., Kosoy, O., Laven, J., Panella, A., Velez, J., Lambert, A.J., et al., 2006. Chikungunya Virus in US Travelers Returning from India, vol. 13 Number 5—May 2007 - Emerging Infectious Disease Journal - CDC. [cited 2018 Feb 19]; Available from: https://wwwnc.cdc.gov/eid/article/13/5/07-0015_article.
- Laurent, P., Roux, K.L., Grivard, P., Bertil, G., Naze, F., Picard, M., et al., 2007 Aug 1. Development of a sensitive real-time reverse transcriptase PCR assay with an internal control to detect and quantify chikungunya virus. *Clin. Chem.* 53 (8), 1408–1414.
- Lednický, J., De Rochars, V.M.B., Elbadry, M., Loeb, J., Telisma, T., Chavannes, S., et al., 2016. Mayaro virus in child with acute febrile illness, Haiti, 2015. *Emerg. Infect. Dis.* 22 (11) 2000–2.
- LeDuc, J.W., Pinheiro, F.P., Rosa, APAT da, 1981 May 1. An outbreak of Mayaro virus disease in Belterra, Brazil. *Am. J. Trop. Med. Hyg.* 30 (3), 682–688.
- Lemant, J., Boisson, V., Winer, A., Thibault, L., André, H., Tixier, F., et al., 2008. Serious acute chikungunya virus infection requiring intensive care during the reunion island outbreak in 2005–2006*. *Crit. Care Med.* 36 (9), 2536–2541.
- Leo, Y.S., Chow, A.L.P., Tan, L.K., Lye, D.C., Lin, L., Ng, L.C., 2009 May. Chikungunya outbreak, Singapore, 2008. *Emerg. Infect. Dis.* 15 (5), 836–837.
- Lepar-Goffart, I., Nougairède, A., Cassadou, S., Prat, C., De Lamballerie, X., 2014. Chikungunya in the Americas. *Lancet* 383 (9916), 514.
- Litzba, N., Schuffenecker, I., Zeller, H., Drosten, C., Emmerich, P., Charrel, R., et al., 2008. Evaluation of the first commercial chikungunya virus indirect immunofluorescence test. *J. Virol Methods* 149 (1), 175–179.
- Liu, J., Ochieng, C., Wiersma, S., Ströher, U., Townner, J.S., Whitmer, S., et al., 2016 Jan. Development of a TaqMan array card for acute-febrile-illness outbreak investigation and surveillance of emerging pathogens, including ebola virus. *J. Clin. Microbiol.* 54 (1), 49–58.
- Llagonne-Barets, M., Icard, V., Lepar-Goffart, I., Prat, C., Perpoint, T., André, P., et al., 2016 Apr 1. A case of Mayaro virus infection imported from French Guiana. *J. Clin. Virol.* 77 (Suppl. C), 66–68.
- Long, K.C., Ziegler, S.A., Thangamani, S., Hausser, N.L., Kochel, T.J., Higgs, S., et al., 2011 Oct. Experimental transmission of Mayaro virus by *Aedes aegypti*. *Am. J. Trop. Med. Hyg.* 85 (4), 750–757.
- Lu, X., Li, X., Mo, Z., Jin, F., Wang, B., Zhao, H., et al., 2012 Nov. Rapid identification of Chikungunya and Dengue virus by a real-time reverse transcription-loop-mediated isothermal amplification method. *Am. J. Trop. Med. Hyg.* 87 (5), 947–953.
- Lwande, O.W., Obanda, V., Bucht, G., Mosomtai, G., Otieno, V., Ahm, C., et al., 2015 Dec 18. Global emergence of Alphaviruses that cause arthritis in humans. *Infect. Ecol. Epidemiol.* 5 [Internet] [cited 2018 Nov 5]. Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC4685977/>.
- Mackay, I.M., Arden, K.E., 2016 Dec 1. Mayaro virus: a forest virus primed for a trip to the city? *Microb. Infect.* 18 (12), 724–734.
- Malvy, D., Ezzedine, K., Mamani-Matsuda, M., Autran, B., Tolou, H., Receveur, M.-C., et al., 2009. Destructive arthritis in a patient with chikungunya virus infection with persistent specific IgM antibodies. *BMC Infect. Dis.* 9, 200.
- Mead, P.S., Duggal, N.K., Hook, S.A., Delorey, M., Fischer, M., Olzenak McGuire, D., et al., 2018 Apr 12. Zika virus shedding in semen of symptomatic infected men. *N. Engl. J. Med.* 378 (15), 1377–1385.
- Mehta, R., Soares, C.N., Medialdea-Carrera, R., Ellul, M., da Silva, M.T.T., Rosala-Hallas, A., et al., 2018 Feb 12. The spectrum of neurological disease associated with Zika and chikungunya viruses in adults in Rio de Janeiro, Brazil: a case series. *PLoS Negl. Trop. Dis.* 12 (2) [Internet] [cited 2018 Dec 28]; Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC5837186/>.
- Mota, M.T. de O., Ribeiro, M.R., Vedovello, D., Nogueira, M.L., 2015 Sep 1. Mayaro virus: a neglected arbovirus of the Americas. *Future Virol.* 10 (9), 1109–1122.
- Musso, D., Teissier, A., Rouault, E., Teururai, S., de Pina, J.-J., Nhan, T.-X., 2016. Detection of chikungunya virus in saliva and urine. *Virol. J.* 13 (1), 102.
- Nagpal, K., Agarwal, P., Kumar, A., Reddi, R., 2017. Chikungunya infection presenting as mild encephalitis with a reversible lesion in the splenium: a case report. *J. Neurovirol.* 23 (3), 501–503.
- Naveca, F.G., Nascimento, VA do, Souza, VC de, Nunes, B.T.D., Rodrigues, D.S.G., Vasconcelos, P.F. da C., et al., 2017 Jul. Multiplexed reverse transcription real-time polymerase chain reaction for simultaneous detection of Mayaro, Oropouche, and Oropouche-like viruses. *Mem. Inst. Oswaldo Cruz* 112 (7), 510–513.
- Niedrig, M., Zeller, H., Schuffenecker, I., Drosten, C., Emmerich, P., Rumer, L., et al., 2009 Sep 1. International diagnostic accuracy study for the serological detection of chikungunya virus infection. *Clin. Microbiol. Infect.* 15 (9), 880–884.
- Nitatattana, N., Kanjanapas, K., Yoksan, S., Satimai, W., Vongba, N., Langdatsuwana, S., et al., 2014 Oct 21. Long-term persistence of Chikungunya virus neutralizing antibodies in human populations of North Eastern Thailand. *Virol. J.* 11 [Internet] [cited 2018 Feb 26]. Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC4283153/>.
- Oliver, M., Grandadam, M., Marimoutou, C., Rogier, C., Botelho-Nevers, E., Tolou, H., et al., 2009 Feb 3. Persisting mixed cryoglobulinemia in chikungunya infection. *PLoS Negl. Trop. Dis.* 3 (2) [Internet] [cited 2018 Jan 29]. Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC2629124/>.
- PAHO WHO|Chikungunya|Data, Maps and Statistics [Internet]. [cited 2019 Jan 28].

- Available from: https://www.paho.org/hq/index.php?option=com_topics&view=rdmore&cid=5927&item=chikungunya&type=statistics&Itemid=40931&lang=en.
- Panning, M., Grywna, K., van Esbroeck, M., Emmerich, P., Drosten, C., 2008 Mar. Chikungunya fever in travelers returning to Europe from the Indian Ocean region, 2006. *Emerg. Infect. Dis.* 14 (3), 416–422.
- Panning, M., Hess, M., Fischer, W., Grywna, K., Pfeiffer, M., Drosten, C., 2009 Sep. Performance of the RealStar chikungunya virus real-time reverse transcription-PCR kit. *J. Clin. Microbiol.* 47 (9), 3014–3016.
- Panning, M., Charrel, R.N., Mantke, O.D., Landt, O., Niedrig, M., Drosten, C., 2009b. Coordinated implementation of chikungunya virus reverse transcription-PCR. *Emerg. Infect. Dis.* 15 (3), 469–471.
- Parida, M.M., Santhosh, S.R., Dash, P.K., Tripathi, N.K., Lakshmi, V., Mamidi, N., et al., 2007. Rapid and real-time detection of Chikungunya virus by reverse transcription loop-mediated isothermal amplification assay. *J. Clin. Microbiol.* 45 (2), 351–357.
- Partidos, C.D., Paykel, J., Weger, J., Borland, E.M., Powers, A.M., Seymour, R., et al., 2012 Jun 29. Cross-protective immunity against O'nyong-nyong virus afforded by a novel recombinant chikungunya vaccine. *Vaccine* 30 (31), 4638–4643.
- Pastorino, B., Bessaud, M., Grandadam, M., Murri, S., Tolou, H.J., Peyrefitte, C.N., 2005. Development of a TaqMan® RT-PCR assay without RNA extraction step for the detection and quantification of African Chikungunya viruses. *J. Virol Methods* 124 (1–2), 65–71.
- PEI, Paul-Ehrlich-Institut: Information on WHO Biological Reference Preparation, Chikungunya virus RNA for NAT-based assays [Internet] [cited 2018 Feb 19]. Available from: <https://www.pei.de/EN/information/license-applicants/standard-and-referencematerials/who/chikungunya-virus-rna/chikungunya-node.html>.
- Porterfield, J.S., 1961. Cross-neutralization studies with group A arthropod-borne viruses. *Bull. World Health Organ.* 24, 735–741.
- Powers, A.M., Logue, C.H., 2007 Sep. Changing patterns of chikungunya virus: re-emergence of a zoonotic arbovirus. *J. Gen. Virol.* 88 (Pt 9), 2363–2377.
- Powers, A.M., Brault, A.C., Shirako, Y., Strauss, E.G., Kang, W., Strauss, J.H., et al., 2001 Nov. Evolutionary relationships and systematics of the alphaviruses. *J. Virol.* 75 (21), 10118–10131.
- Prat, C.M., Flusin, O., Panella, A., Tenebray, B., Lanciotti, R., Leparco-Goffart, I., 2014 Dec. Evaluation of commercially available serologic diagnostic tests for chikungunya virus. *Emerg. Infect. Dis.* 20 (12), 2129–2132.
- Prince, H.E., Seaton, B.L., Matud, J.L., Batterman, H.J., 2015 Jan 3. Chikungunya virus RNA and antibody testing at a national reference laboratory since the emergence of chikungunya virus in the Americas. *Clin. Vaccine Immunol.* 22 (3), 291–297.
- Prince, H.E., Altrich, M.L., Nowicki, M.J., 2016 Jan 10. Evaluation of two enzyme-linked immunosorbent assay kits for chikungunya virus IgM using samples from deceased organ and tissue donors. *Clin. Vaccine Immunol.* 23 (10), 825–830.
- Pyndiah, M.N., Pursem, V., Meeto, G., Daby, S., Ramuth, V., Binkah, P., et al., 2012 Mar. Chikungunya virus isolation using simplified cell culture technique in Mauritius. *Med. Trop. Rev. Corps Sante Colon* 72 Spec No:63–5.
- RealStar Chikungunya RT-PCR Kit 2.0_WEB_CE_EN-S02.Pdf [Internet] [cited 2018 Nov 5]. Available from: http://www.altona-diagnostics.com/files/public/Content%20Homepage/%2002%20RealStar/MAN%20-%20CE%20-%20EN/RealStar%20Chikungunya%20RT-PCR%20Kit%202.0_WEB_CE_EN-S02.pdf.
- Receveur, M.C., Grandadam, M., Pistone, T., Malvy, D., 2010 May 6. Infection with Mayaro virus in a French traveller returning from the Amazon region, Brazil, January, 2010. *Euro Surveill.* 15 (18), 19563.
- Reddy, V., Ravi, V., Desai, A., Parida, M., Powers, A.M., Johnson, B.W., 2012 Nov 1. Utility of IgM ELISA, TaqMan real-time PCR, reverse transcription PCR, and RT-LAMP assay for the diagnosis of Chikungunya fever. *J. Med. Virol.* 84 (11), 1771–1778.
- Research C for BE and. Blood & Blood Products [Internet] [cited 2018 Feb 20]. Available from: <https://www.fda.gov/BiologicsBloodVaccines/BloodBloodProducts/default.htm>.
- Rezza, G., Chen, R., Weaver, S.C., 2017. O'nyong-nyong fever: a neglected mosquito-borne viral disease. *Pathog. Glob. Health* 7724 (October), 1–5.
- Rocha, V.F.D., de Oliveira, A.H.P., Bandeira, A.C., Sardi, S.I., Garcia, R.F., Magalhães, S. de A., et al., 2018. Chikungunya virus infection associated with encephalitis and anterior uveitis. *Ocul. Immunol. Inflamm.* 26 (5), 677–679.
- Rwaguma, E.B., Lutwama, J.J., Sempala, S.D., Kiwanuka, N., Kamugisha, J., Okware, S., et al., 1997. Emergence of epidemic O'nyong-nyong fever in southwestern Uganda, after an absence of 35 years. *Emerg. Infect. Dis.* 3 (1), 77.
- Saxton-Shaw, K.D., Ledermann, J.P., Borland, E.M., Stovall, J.L., Mossel, E.C., Singh, A.J., et al., 2013. O'nyong nyong virus molecular determinants of unique vector specificity reside in non-structural protein 3. *PLoS Negl. Trop. Dis.* 7 (1).
- Schmidt, J.R., Gajdusek, D.C., Schaeffer, M., Gorrie, R.H., 1959 Jul 1. Epidemic jungle fever among okinawan colonists in the Bolivian rain forest. *Am. J. Trop. Med. Hyg.* 8 (4), 479–487.
- Schuffenecker, I., Itean, I., Michault, A., Murri, S., Frangeul, L., Vaney, M.-C., et al., 2006 May 23. Genome microevolution of chikungunya viruses causing the Indian ocean outbreak. *PLoS Med.* 3 (7), e263.
- Simmons, G., Brès, V., Lu, K., Liss, N.M., Brambilla, D.J., Ryff, K.R., et al., 2016. High incidence of chikungunya virus and frequency of viremic blood donations during epidemic, Puerto Rico, USA, 2014. *Emerg. Infect. Dis.* 22 (7), 1221–1228.
- Smith, G.C., Franc, D.B., 1991 Mar. Laboratory studies of a Brazilian strain of Aedes albopictus as a potential vector of Mayaro and Oropouche viruses. *J. Am. Mosq. Control Assoc.* 7 (1), 89–93.
- Smith, D.R., Lee, J.S., Jahrling, J., Kulesh, D.A., Turell, M.J., Groebner, J.L., et al., 2009. Development of field-based real-time reverse transcription-polymerase chain reaction assays for detection of Chikungunya and O'nyong-nyong viruses in mosquitoes. *Am. J. Trop. Med. Hyg.* 81 (4), 679–684.
- Smith, J.L., Pugh, C.L., Cisney, E.D., Keasey, S.L., Guevara, C., Ampuero, J.S., et al., 2018. Human antibody responses to emerging Mayaro virus and cocirculating alphavirus infections examined by using structural proteins from nine new and old World lineages. *mSphere* 3 (2) [Internet] Mar 21 [cited 2018 Apr 4]. Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC5863033/>.
- Soh, L.T., Squires, R.C., Tan, L.K., Pok, K.Y., Yang, H., Liew, C., et al., 2016 Apr 22. External quality assessment of dengue and chikungunya diagnostics in the Asia Pacific region, 2015. *West Pac Surveill. Response J. WPSAR* 7 (2), 26–34.
- Sow, A., Faye, O., Diallo, M., Diallo, D., Chen, R., Faye, O., et al., 2017 Dec 2. Chikungunya outbreak in Kedougou, Southeastern Senegal in 2009–2010. *Open Forum Infect. Dis.* 5 (1) [Internet] [cited 2018 Dec 10]. Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC5767945/>.
- Staples, J.E., Breiman, R.F., Powers, A.M., 2009. Chikungunya fever: an epidemiological review of a Re-emerging infectious disease. *Clin. Infect. Dis.* 49 (6), 942–948.
- Strauss, J.H., Strauss, E.G., 1994. The alphaviruses: gene expression, replication, and evolution. *Microbiol. Rev.* 58 (3), 491–562.
- Tappe, D., Kapaun, A., Emmerich, P., Campos, R. de M., Cadar, D., Günther, S., et al., 2014 Oct. O'nyong-nyong virus infection imported to Europe from Kenya by a traveler. *Emerg. Infect. Dis.* 20 (10), 1766–1767.
- Taraphdar, D., Roy, B.K., Chatterjee, S., 2015 Feb. Chikungunya virus infection amongst the acute encephalitis syndrome cases in West Bengal, India. *Indian J. Med. Microbiol.* 33 (Suppl. 1), 153–156.
- Thiberville, S.-D., Moyn, N., Dupuis-Maguiraga, L., Nougaiere, A., Gould, E.A., Roques, P., et al., 2013a. Chikungunya fever: epidemiology, clinical syndrome, pathogenesis and therapy. *Antivir. Res.* 99 (3), 345–370.
- Thiberville, S.-D., Boisson, V., Gaudart, J., Simon, F., Flahault, A., de Lamballerie, X., 2013 Jan 17b. Chikungunya fever: a clinical and virological investigation of outpatients on reunion island, South-West Indian Ocean. *PLoS Negl. Trop. Dis.* 7 (1) [Internet] [cited 2018 May 18]. Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC3547841/>.
- Tournebise, P., Charlin, C., Lagrange, M., 2009 Jan. [Neurological manifestations in chikungunya: about 23 cases collected in reunion island]. *Rev. Neurol. (Paris)* 165 (1), 48–51.
- Tssetsarkin, K.A., Vanlandingham, D.L., McGee, C.E., Higgs, S., 2007 Dec 7. A single mutation in chikungunya virus affects vector specificity and epidemic potential. *PLoS Pathog.* 3 (12) e201.
- Tssetsarkin, K.A., Chen, R., Weaver, S.C., 2016 Feb. Interspecies transmission and chikungunya virus emergence. *Curr. Opin. Virol.* 16, 143–150.
- Vanlandingham, D.L., Hong, C., Klingler, K., Tssetsarkin, K., McElroy, K.L., Powers, A.M., et al., 2005. Differential infectivities of O'nyong-nyong and chikungunya virus isolates in Anopheles gambiae and Aedes aegypti mosquitoes. *Am. J. Trop. Med. Hyg.* 72 (5), 616–621.
- Vega-Rua, A., Zouache, K., Girod, R., Failloux, A.-B., Lourenco-de-Oliveira, R., 2014. High level of vector competence of Aedes aegypti and Aedes albopictus from ten American countries as a crucial factor in the spread of chikungunya virus. *J. Virol.* 88 (11), 6294–6306.
- Verma, P., Bhatnagar, S., Kumar, P., Chattree, V., Parida, M.M., Hoti, S.L., et al., 2013. Analysis of antibody response (IgM, IgG, IgG3) to Chikungunya virus using panel of peptides derived from envelope protein for serodiagnosis. *Clin. Chem. Lab. Med. CCLM* 52 (2), 297–307.
- Virus Pathogen Database and Analysis Resource (ViPR) - Genome Database with Visualization and Analysis Tools [Internet] [cited 2019 Feb 25]. Available from: <https://www.viprbrc.org/brc/home.spg?decorator=vipr>.
- Virus Variation [Internet] [cited 2019 Feb 25]. Available from: <https://www.ncbi.nlm.nih.gov/genome/viruses/variation/>.
- Waggoner, J.J., Ballesteros, G., Gresh, L., Mohamed-Hadley, A., Tellez, Y., Sahoo, M.K., et al., 2016 May. Clinical evaluation of a single-reaction real-time RT-PCR for dengue and chikungunya virus detection. *J. Clin. Virol. Off. Publ. Pan Am. Soc. Clin. Virol.* 78, 57–61.
- Waggoner, J.J., Heath, C.J., Ndenga, B., Mutuku, F., Sahoo, M.K., Mohamed-Hadley, A., et al., 2017 Jul 12a. Development of a real-time reverse transcription polymerase chain reaction for O'nyong-nyong virus and evaluation with clinical and mosquito specimens from Kenya. *Am. J. Trop. Med. Hyg.* 97 (1), 121–124.
- Waggoner, J.J., Rojas, A., Mohamed-Hadley, A., de Guillén, Y.A., Pinsky, B.A., 2017 Nov 21b. Real-time RT-PCR for Mayaro virus detection in plasma and urine. *J. Clin. Virol. Off. Publ. Pan Am. Soc. Clin. Virol.* 98, 1–4.
- WER9033_410-414.pdf [Internet] [cited 2018 Nov 14]. Available from: http://apps.who.int/iris/bitstream/handle/10665/242406/WER9033_410-414.PDF;jsessionid=8BAAD8AC72C8B24DC2BFBB9920686F7?sequence=1.
- who-collaborating-study-chikungunya.pdf [Internet] [cited 2018 Feb 19]. Available from: https://www.pei.de/SharedDocs/Downloads/EN/who/who-collaborating-study-chikungunya.pdf?__blob=publicationFile&v=2.
- Williams, M.C., Woodall, J.P., Corbet, P.S., Gillett, J.D., 1965 May. O'nyong-nyong fever: an epidemic virus disease in east africa. 8. virus isolations from anopheles mosquitoes. *Trans. R. Soc. Trop. Med. Hyg.* 59, 300–306.
- Win, M.K., Chow, A., Dimataca, F., Go, C.J., Leo, Y.S., 2010 Oct 1. Chikungunya fever in Singapore: acute clinical and laboratory features, and factors associated with persistent arthralgia. *J. Clin. Virol.* 49 (2), 111–114.
- Yap, G., Pok, K.-Y., Lai, Y.-L., Hapuarachchi, H.-C., Chow, A., Leo, Y.-S., et al., 2010 Jul 20. Evaluation of chikungunya diagnostic assays: differences in sensitivity of serology assays in two independent outbreaks. *PLoS Negl. Trop. Dis.* 4 (7) [Internet] [cited 2018 Feb 2]. Available from: <https://www.ncbi.nlm.nih.gov/pmc/articles/PMC2907414/>.